

- ID label can not be attached to the probe housing as it comes into contact with aggressive chemical liquids.
- The ID label will be attached on cable end (close to the plug). The ID label will be not removable.
- FCC statement does not fit on the ID label and comes instead in the manual