### DASY/EASY - Parameters of Probe: EX3DV4 - SN:3958

### **Basic Calibration Parameters**

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm $(\mu V/(V/m)^2)^A$	0.50	0.45	0.52	± 10.1 %
DCP (mV) <sup>B</sup>	98.6	102.2	97.4	** ; . ; * * * * * *

#### **Modulation Calibration Parameters**

UID	Communication System Name		A dB	B dB√μV	С	D dB	VR mV	Unc <sup>E</sup> (k=2)
0	CW	Х	0.0	0.0	1.0	0.00	143.6	±3.0 %
		Υ	0.0	0.0	1.0		158.6	
		Z	0.0	0.0	1.0		145.2	

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>B</sup> Numerical linearization parameter: uncertainty not required.

A The uncertainties of Norm X,Y,Z do not affect the E2-field uncertainty inside TSL (see Pages 5 and 6).

E Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

EX3DV4- SN:3958 July 26, 2016

### DASY/EASY - Parameters of Probe: EX3DV4 - SN:3958

### Calibration Parameter Determined in Head Tissue Simulating Media

f (MHz) <sup>C</sup>	Relative Permittivity <sup>F</sup>	Conductivity (S/m) F	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k=2)	
750	41.9	0.89	10.85	10.85	10.85	0.59	0.80	± 12.0 %	
835	41.5	0.90	10.62	10.62	10.62	0.49	0.80	± 12.0 %	
900	41.5	0.97	10.33	10.33	10.33	0.27	1.19	± 12.0 %	
1450	40.5	1.20	9.21	9.21	9.21	0.36	0.80	± 12.0 %	
1750	40.1	1.37	8.82	8.82	8.82	0.42	0.80	± 12.0 %	
1900	40.0	1.40	8.49	8.49	8.49	0.44	0.80	± 12.0 %	
2000	40.0	1.40	8.58	8.58	8.58	0.39	0.80	± 12.0 %	
2300	39.5	1.67	8.15	8.15	8.15	0.44	0.80	± 12.0 %	
2450	39.2	1.80	7.84	7.84	7.84	0.38	0.90	± 12.0 %	
2600	39.0	1.96	7.69	7.69	7.69	0.38	0.93	± 12.0 %	
3500	37.9	2.91	7.30	7.30	7.30	0.35	1.10	± 13.1 %	
5200	36.0	4.66	5.88	5.88	5.88	0.35	1.80	± 13.1 %	
5600	35.5	5.07	5.20	5.20	5.20	0.40	1.80	± 13.1 %	
5750	35.4	5.22	5.38	5.38	5.38	0.40	1.80	± 13.1 %	

 $<sup>^{\</sup>rm C}$  Frequency validity above 300 MHz of  $\pm$  100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to  $\pm$  50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is  $\pm$  10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to  $\pm$  110 MHz.

validity can be extended to ± 110 MHz.

F At frequencies below 3 GHz, the validity of tissue parameters (ε and σ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ε and σ) is restricted to ± 5%. The uncertainty is the RSS of the CopyE uncertainty for indicated target tissue parameters.

the ConvF uncertainty for indicated target tissue parameters.

Galpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

## DASY/EASY - Parameters of Probe: EX3DV4 - SN:3958

### Calibration Parameter Determined in Body Tissue Simulating Media

f (MHz) <sup>C</sup>	Relative Permittivity <sup>F</sup>			ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k=2)	
750 55.5		0.96	10.29	10.29	10.29	0.49		± 12.0 %	
835	55.2	0.97	10.34	10.34	10.34	0.43	0.85	± 12.0 %	
1750	53.4	1.49	8.58	8.58	8.58	0.38	0.80	± 12.0 %	
1900	53.3	1.52	8.01	8.01	8.01	0.32	0.94	± 12.0 %	
2300	52.9	1.81	8.02	8.02	8.02	0.37	0.80	± 12.0 %	
2450	52.7	1.95	7.72	7.72	7.72	0.42	0.80	± 12.0 %	
2600	52.5	2.16	7.62	7.62	7.62	0.36	0.80	± 12.0 %	
3500	51.3	3.31	7.10	7.10	7.10	0.30	1.20	± 13.1 %	
5250	48.9	5.36	4.46	4.46	4.46	0.45	1.90	± 13.1 %	
5600	48.5	5.77	3.76	3.76	3.76	0.55	1.90	± 13.1 %	
5750	48.3	5.94	4.03	4.03	4.03	0.55	1.90	± 13.1 %	

<sup>&</sup>lt;sup>c</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.


At frequencies below 2 CHz, the weight in the convergence of the

F At frequencies below 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) can be relaxed to  $\pm$  10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters ( $\epsilon$  and  $\sigma$ ) is restricted to  $\pm$  5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

G Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is

Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

# Frequency Response of E-Field (TEM-Cell:ifi110 EXX, Waveguide: R22)





Uncertainty of Frequency Response of E-field: ± 6.3% (k=2)


# Receiving Pattern ( $\phi$ ), $\vartheta = 0^{\circ}$

f=600 MHz,TEM

f=1800 MHz,R22









Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)


## Dynamic Range f(SAR<sub>head</sub>) (TEM cell , f<sub>eval</sub>= 1900 MHz)






Uncertainty of Linearity Assessment: ± 0.6% (k=2)

## **Conversion Factor Assessment**



# **Deviation from Isotropy in Liquid**

Error  $(\phi, \vartheta)$ , f = 900 MHz



0.0

Uncertainty of Spherical Isotropy Assessment: ± 2.6% (k=2)

0.2

0.4

0.6

-0.6 -0.4

-0.2

# DASY/EASY - Parameters of Probe: EX3DV4 - SN:3958

### **Other Probe Parameters**

Sensor Arrangement						. 1111 ( )			Triangular
Connector Angle (°)			*	- 1 - 1		144, 1	* * * * * *		41
Mechanical Surface Detection Mode					1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	*.			enabled
Optical Surface Detection Mode	· · ·	: 1	****			*****			disabled
Probe Overall Length	N. N.								337 mm
Probe Body Diameter		:				······································		÷ .	10 mm
Tip Length	1	\$	;						9 mm
Tip Diameter					1				2.5 mm
Probe Tip to Sensor X Calibration Point	-								1 mm
Probe Tip to Sensor Y Calibration Point	1 1 1								1 mm
Probe Tip to Sensor Z Calibration Point									1 mm
Recommended Measurement Distance f	rom	Surfac	e	W-21-1					1.4 mm