



77 Progress Parkway  
St. Louis, MO 63043

Tel. 314.576.6800  
Fax 877.581.7858

[www.biomedsys.com](http://www.biomedsys.com)

---

**QUALITY MANAGEMENT**

30 March 2012

RE: FCC ID: YCVBRHA02

The only change in the device under application for FCC ID YCVBRHA02 from the device previously approved under FCC ID YCVBRHA01 is an upgrade of the Sierra Wireless GSM module from the Q2687 G to the Q2687 RD. The Bluetooth section is therefore identical to the device approved under FCC ID YCVBRHA01.

A handwritten signature in black ink, appearing to read "K. Michael Kroehnke".

K. Michael Kroehnke  
Director, Quality Management