

Appendix C for KSCR221100223601

Calibration Certificate

| Object | Apply | No | Model | SN | Calibration Date |
|--------|-------------------------------------|----|---------|-------|------------------|
| Dipole | <input type="checkbox"/> | 1 | CLA150 | 4025 | 2021/04/26 |
| | <input type="checkbox"/> | 2 | D450V3 | 1103 | 2021/04/21 |
| | <input type="checkbox"/> | 3 | D750V3 | 1188 | 2022/03/29 |
| | <input type="checkbox"/> | 4 | D835V2 | 4d114 | 2022/03/31 |
| | <input type="checkbox"/> | 5 | D900V2 | 1d079 | 2022/06/07 |
| | <input type="checkbox"/> | 6 | D1800V2 | 2d170 | 2022/03/31 |
| | <input type="checkbox"/> | 7 | D1900V2 | 5d136 | 2022/06/07 |
| | <input type="checkbox"/> | 8 | D2000V2 | 1041 | 2022/06/06 |
| | <input type="checkbox"/> | 9 | D2300V2 | 1096 | 2022/03/31 |
| | <input checked="" type="checkbox"/> | 10 | D2450V2 | 817 | 2022/04/01 |
| | <input type="checkbox"/> | 11 | D2600V2 | 1158 | 2022/03/31 |
| | <input checked="" type="checkbox"/> | 12 | D5GHzV2 | 1095 | 2022/06/01 |
| DAE | <input checked="" type="checkbox"/> | 13 | DAE4 | 1245 | 2022/05/30 |
| Probe | <input checked="" type="checkbox"/> | 14 | EX3DV4 | 7346 | 2022/03/30 |



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com
 No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1 Dipole

1.1 CLA150 - SN 4025

| <p>Calibration Laboratory of Schmid & Partner Engineering AG Zehrwegstrasse 43, 8004 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Client: SGS-CN (Auden) Certificate No.: CLA150-4025_Apr21</p> <p>CALIBRATION CERTIFICATE</p> <p>Object: CLA150 - SN: 4025</p> <p>Calibration procedure(s): QA CAL-15-V9 Calibration Procedure for SAR Validation Sources below 700 MHz</p> <p>Calibration date: April 26, 2021</p> <p>The calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility; environment temperature (22 ± 3)°C and humidity < 70%.</p> <p>Calibration Equipment used (MPE critical for calibration)</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power meter NRP</td> <td>SN: 10476</td> <td>09-Apr-21 (No. 217-939110320)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NRP Z91</td> <td>SN: 10384</td> <td>09-Apr-21 (No. 217-93521)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor NRP Z91</td> <td>SN: 10385</td> <td>09-Apr-21 (No. 217-93522)</td> <td>Apr-22</td> </tr> <tr> <td>Reference 20 dB Attenuator</td> <td>SN: C22862 (230)</td> <td>09-Apr-21 (No. 217-93343)</td> <td>Apr-22</td> </tr> <tr> <td>Type-N mismatch combination</td> <td>SN: 31982 / 0037</td> <td>09-Apr-21 (No. 217-93344)</td> <td>Apr-22</td> </tr> <tr> <td>Reference Probe EX3004 (DIE4)</td> <td>SN: 3877</td> <td>30-Dec-20 (No. EX3-3877_Dec20)</td> <td>Dec-21</td> </tr> <tr> <td></td> <td>SN: 664</td> <td>25-Jan-20 (No. DMS4-656_Jan20)</td> <td>Jan-21</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Check Date (in house)</th> <th>Scheduled Check</th> </tr> </thead> <tbody> <tr> <td>Power meter E4419B</td> <td>SN: G814388214</td> <td>08-Apr-16 (in house check Jun-20)</td> <td>In house check Jun-22</td> </tr> <tr> <td>Power sensor E4412A</td> <td>SN: M41488067</td> <td>06-Apr-16 (in house check Jun-20)</td> <td>In house check Jun-22</td> </tr> <tr> <td>Power sensor E4412A</td> <td>SN: 00010210</td> <td>06-Apr-16 (in house check Jun-20)</td> <td>In house check Jun-22</td> </tr> <tr> <td>RF generator HP 85840</td> <td>SN: US2400107109</td> <td>04-Aug-09 (in house check Jun-20)</td> <td>In house check Jun-22</td> </tr> <tr> <td>Network Analyzer Agilent E8363A</td> <td>SN: US44000477</td> <td>31-Mar-14 (in house check Oct-20)</td> <td>In house check Oct-21</td> </tr> </tbody> </table> <p>Calibrated by: Jeffrey Kitzman Function: Laboratory Technician Signature: <i>[Signature]</i></p> <p>Approved by: Kate Polovic Technical Manager Signature: <i>[Signature]</i></p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory. Issued: April 26, 2021</p> <p>Certificate No: CLA150-4025_Apr21 Page 1 of 6</p> | Primary Standards | ID # | Cal Date (Certificate No.) | Scheduled Calibration | Power meter NRP | SN: 10476 | 09-Apr-21 (No. 217-939110320) | Apr-22 | Power sensor NRP Z91 | SN: 10384 | 09-Apr-21 (No. 217-93521) | Apr-22 | Power sensor NRP Z91 | SN: 10385 | 09-Apr-21 (No. 217-93522) | Apr-22 | Reference 20 dB Attenuator | SN: C22862 (230) | 09-Apr-21 (No. 217-93343) | Apr-22 | Type-N mismatch combination | SN: 31982 / 0037 | 09-Apr-21 (No. 217-93344) | Apr-22 | Reference Probe EX3004 (DIE4) | SN: 3877 | 30-Dec-20 (No. EX3-3877_Dec20) | Dec-21 | | SN: 664 | 25-Jan-20 (No. DMS4-656_Jan20) | Jan-21 | Secondary Standards | ID # | Check Date (in house) | Scheduled Check | Power meter E4419B | SN: G814388214 | 08-Apr-16 (in house check Jun-20) | In house check Jun-22 | Power sensor E4412A | SN: M41488067 | 06-Apr-16 (in house check Jun-20) | In house check Jun-22 | Power sensor E4412A | SN: 00010210 | 06-Apr-16 (in house check Jun-20) | In house check Jun-22 | RF generator HP 85840 | SN: US2400107109 | 04-Aug-09 (in house check Jun-20) | In house check Jun-22 | Network Analyzer Agilent E8363A | SN: US44000477 | 31-Mar-14 (in house check Oct-20) | In house check Oct-21 | <p>Calibration Laboratory of Schmid & Partner Engineering AG Zehrwegstrasse 43, 8004 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Accreditation No.: SCS 0108</p> <p>Glossary:</p> <p>TSL: Issue simulating liquid sensitivity in TSL / NORM x,y,z</p> <p>ConnF: not applicable or not measured</p> <p>N/A: not applicable or not measured</p> <p>Calibration is Performed According to the Following Standards:</p> <ol style="list-style-type: none"> IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013 IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2016 IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010 KDB 865864, "SAR Measurement Requirements for 100 MHz to 6 GHz" <p>Additional Documentation:</p> <ol style="list-style-type: none"> DASY4/5 System Handbook <p>Methods Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in this certificate are valid at the frequency indicated. Antenna Parameters with TSL: The source is mounted in a touch configuration below the center marking of the flat phantom. Return Loss: This parameter is measured with the source positioned under the liquid filled phantom (as described in the measurement condition clause). The Return Loss ensures low reflected power. No uncertainty required. SAR measured: SAR measured at the stated antenna input power. SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. <p>The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by its coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.</p> <p>Certificate No: CLA150-4025_Apr21 Page 2 of 6</p> | | |
|---|------------------------------|-----------------------------------|----------------------------|-----------------------|------------------------|-----------|-------------------------------|------------------|-----------------------------|-----------------|---------------------------|--------|----------------------|------------------------------|---------------------------------|-----------|----------------------------|------------------|---------------------------|-------------|-----------------------------|------------------|-----------------------------|---------|-------------------------------|-----------|--------------------------------|-----------------|------------|-----------------|---|----------|---------------------|------|---|-----------------|--------------------|----------------|-----------------------------------|-----------------------|-------------------------------------|------------------|-----------------------------------|---|---------------------|--------------|-----------------------------------|-----------------------|-----------------------|-------------------------------------|-----------------------------------|---------------------------------|--|--------------------------------------|-----------------------------------|-----------------------|---|-----------------|-------|
| Primary Standards | ID # | Cal Date (Certificate No.) | Scheduled Calibration | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power meter NRP | SN: 10476 | 09-Apr-21 (No. 217-939110320) | Apr-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor NRP Z91 | SN: 10384 | 09-Apr-21 (No. 217-93521) | Apr-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor NRP Z91 | SN: 10385 | 09-Apr-21 (No. 217-93522) | Apr-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Reference 20 dB Attenuator | SN: C22862 (230) | 09-Apr-21 (No. 217-93343) | Apr-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Type-N mismatch combination | SN: 31982 / 0037 | 09-Apr-21 (No. 217-93344) | Apr-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Reference Probe EX3004 (DIE4) | SN: 3877 | 30-Dec-20 (No. EX3-3877_Dec20) | Dec-21 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| | SN: 664 | 25-Jan-20 (No. DMS4-656_Jan20) | Jan-21 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Secondary Standards | ID # | Check Date (in house) | Scheduled Check | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power meter E4419B | SN: G814388214 | 08-Apr-16 (in house check Jun-20) | In house check Jun-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor E4412A | SN: M41488067 | 06-Apr-16 (in house check Jun-20) | In house check Jun-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor E4412A | SN: 00010210 | 06-Apr-16 (in house check Jun-20) | In house check Jun-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| RF generator HP 85840 | SN: US2400107109 | 04-Aug-09 (in house check Jun-20) | In house check Jun-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Network Analyzer Agilent E8363A | SN: US44000477 | 31-Mar-14 (in house check Oct-20) | In house check Oct-21 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>Measurement Conditions</p> <p>DASY system configuration, as far as not given on page 1:</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>DASY Version</th> <th>Advanced Estimation</th> <th>V32.16.4</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>ELN Flat Phantom</td> <td>Shell thickness: 2 ± 0.2 mm</td> </tr> <tr> <td>EUT Positioning</td> <td>Touch Position</td> <td></td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy = 4.0 mm, dz = 1.4 mm</td> <td>Grade Ratio = 1.4 (Z direction)</td> </tr> <tr> <td>Frequency</td> <td>156 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <p>Head TSL parameters</p> <p>The following parameters and calculations were applied:</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Parameter</th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>62.3</td> <td>0.75 mS/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>51.1 ± 6 %</td> <td>0.75 mS/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td>< 0.5 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p>SAR result with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>SAR averaged over 1 cm³ (1 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>1 W input power</td> <td>3.90 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>3.88 W/kg ± 18.4 % (k=2)</td> </tr> </tbody> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>SAR averaged over 10 cm³ (10 g) of Head TSL</th> <th>condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>1 W input power</td> <td>2.60 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>2.59 W/kg ± 18.0 % (k=2)</td> </tr> </tbody> </table> <p>Certificate No: CLA150-4025_Apr21 Page 3 of 6</p> | DASY Version | Advanced Estimation | V32.16.4 | Extrapolation | Advanced Extrapolation | | Phantom | ELN Flat Phantom | Shell thickness: 2 ± 0.2 mm | EUT Positioning | Touch Position | | Zoom Scan Resolution | dx, dy = 4.0 mm, dz = 1.4 mm | Grade Ratio = 1.4 (Z direction) | Frequency | 156 MHz ± 1 MHz | | Parameter | Temperature | Permittivity | Conductivity | Nominal Head TSL parameters | 22.0 °C | 62.3 | 0.75 mS/m | Measured Head TSL parameters | (22.0 ± 0.2) °C | 51.1 ± 6 % | 0.75 mS/m ± 6 % | Head TSL temperature change during test | < 0.5 °C | --- | --- | SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | | SAR measured | 1 W input power | 3.90 W/kg | SAR for nominal Head TSL parameters | normalized to 1W | 3.88 W/kg ± 18.4 % (k=2) | SAR averaged over 10 cm ³ (10 g) of Head TSL | condition | | SAR measured | 1 W input power | 2.60 W/kg | SAR for nominal Head TSL parameters | normalized to 1W | 2.59 W/kg ± 18.0 % (k=2) | <p>Appendix (Additional assessments outside the scope of SCS 0108)</p> <p>Antenna Parameters with Head TSL</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>47.8 Ω ± 1.5 Ω</td> </tr> <tr> <td>Return Loss</td> <td>-31.4 dB</td> </tr> </tbody> </table> <p>Additional EUT Data</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <tbody> <tr> <td>Manufactured by</td> <td>SPEAQ</td> </tr> </tbody> </table> <p>Certificate No: CLA150-4025_Apr21 Page 4 of 6</p> | Impedance, transformed to feed point | 47.8 Ω ± 1.5 Ω | Return Loss | -31.4 dB | Manufactured by | SPEAQ |
| DASY Version | Advanced Estimation | V32.16.4 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Extrapolation | Advanced Extrapolation | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Phantom | ELN Flat Phantom | Shell thickness: 2 ± 0.2 mm | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| EUT Positioning | Touch Position | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Zoom Scan Resolution | dx, dy = 4.0 mm, dz = 1.4 mm | Grade Ratio = 1.4 (Z direction) | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Frequency | 156 MHz ± 1 MHz | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Parameter | Temperature | Permittivity | Conductivity | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Nominal Head TSL parameters | 22.0 °C | 62.3 | 0.75 mS/m | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 51.1 ± 6 % | 0.75 mS/m ± 6 % | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Head TSL temperature change during test | < 0.5 °C | --- | --- | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SAR measured | 1 W input power | 3.90 W/kg | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SAR for nominal Head TSL parameters | normalized to 1W | 3.88 W/kg ± 18.4 % (k=2) | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SAR averaged over 10 cm ³ (10 g) of Head TSL | condition | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SAR measured | 1 W input power | 2.60 W/kg | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SAR for nominal Head TSL parameters | normalized to 1W | 2.59 W/kg ± 18.0 % (k=2) | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Impedance, transformed to feed point | 47.8 Ω ± 1.5 Ω | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Return Loss | -31.4 dB | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Manufactured by | SPEAQ | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300 t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
 中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

DASY5 Validation Report for Head TSL

Date: 26.04.2021

Test Laboratory: SPEAG, Zurich, Switzerland

DUT: CLA150; Type: CLA150; Serial: CLA150 - SN: 4025

Communication System: UTD 0 - CW; Frequency: 150 MHz
Medium parameters used: $f = 150 \text{ MHz}$; $\sigma = 0.76 \text{ S/m}$; $\epsilon = 51.1$; $\rho = 1000 \text{ kg/m}^3$
Phantom section: Flat Section
Measurement Standard: DASY5 (IEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 - SN3877; ConvF(12.51, 12.51, 12.51) @ 150 MHz; Calibrated: 30.12.2020
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA54 S0654; Calibrated: 26.06.2020
- Phantom: ELL v4.6; Type: QDOVA001BB; Serial: TP.1003
- DASY52 52.10.4(1527); SEMCAD X 14.6.14(7483)

CLA Calibration for HSL-LF Tissue/CLA150, touch configuration, Pin=1W/Zoom Scan, dist=1.4mm (8x10x8)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value $\approx 85.93 \text{ W/m}^2$; Power Drift = -0.02 dB
Peak SAR (extrapolated) = 7.36 W/kg
SAR(1 g) = 3.90 W/kg; SAR(10 g) = 2.60 W/kg
Smallest distance from peaks to all points 3 dB below: Larger than measurement grid (> 30mm)
Ratio of SAR at M2 to SAR at M1 = 80.4%
Maximum value of SAR (measured) = 5.48 W/kg

0 dB = 5.48 W/kg = 7.39 dBW/kg

Certificate No: CLA150-4025_Apr21 Page 6 of 6

Certificate No: CLA150-4025_Apr21 Page 6 of 6

1.2 D450V3 - SN 1103

Calibration Laboratory of Schmid & Partner Engineering AG
Zewgwegstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client: **SGS-CN (Aude)** Certificate No: **D450V3-1103_Apr21**

CALIBRATION CERTIFICATE

Client: **D450V3 - SN: 1103**

Calibrator (procedure): **QA CAL-15_V9**
Calibration Procedure for SAR Validation Sources below 700 MHz

Calibration date: **April 21, 2021**

This calibration certificate documents the possibility to national standards, which define the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility; environment temperature (27 ± 0.1) °C and humidity < 70%.

| Primary Standards | ID # | Exp. Date (Certificate No.) | Scheduled Calibration |
|-----------------------------|--------------------|--------------------------------|-----------------------|
| Power meter NRP | SN: 104778 | 09-Apr-21 (No. 217-0201-09030) | Apr-22 |
| Power sensor NRP-791 | SN: 103344 | 09-Apr-21 (No. 217-0201) | Apr-22 |
| Power sensor NRP-291 | SN: 103345 | 09-Apr-21 (No. 217-0201) | Apr-22 |
| Reference 20 dB Attenuator | SN: C23802 (200) | 09-Apr-21 (No. 217-03043) | Apr-22 |
| Type-N mismatch combination | SN: 3 (0827/06327) | 09-Apr-21 (No. 217-03044) | Apr-22 |
| Reference Probe E3020A | SN: 3077 | 30-Dec-20 (No. E30-3077_Dec20) | Dec-21 |
| DAE4 | SN: 654 | 26-Jun-20 (No. DAE4-654_Jun20) | Jun-21 |

| Secondary Standards | ID # | Check Date (In house) | Scheduled Check |
|---------------------------------|------------------|-----------------------------------|-----------------------|
| Power meter E4419B | SN: G341220274 | 06-Apr-16 (In house check Jun-20) | In house check Jun-21 |
| Power sensor E4412A | SN: EY41496087 | 06-Apr-16 (In house check Jun-20) | In house check Jun-21 |
| Power sensor E4412A | SN: 000110210 | 06-Apr-16 (In house check Jun-20) | In house check Jun-21 |
| HP generator HP 8441C | SN: L52340017100 | 06-Apr-09 (In house check Jun-20) | In house check Jun-22 |
| Network Analyzer Agilent E3036A | SN: L5841980477 | 31-Mar-14 (In house check Oct-20) | In house check Oct-21 |

Calibrated by: **Christof Leuber** (Function: Laboratory Technician)

Approved by: **Kate FORD** (Function: Technical Manager)

Certificate No: D450V3-1103_Apr21 Page 1 of 6

Calibration Laboratory of Schmid & Partner Engineering AG
Zewgwegstrasse 43, 8004 Zurich, Switzerland

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client: **SGS-CN (Aude)** Certificate No: **D450V3-1103_Apr21**

Glossary:

TSL: Issue simulating liquid sensitivity in TSL / NORM x,y,z

ConvF: not applicable or not measured

N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEC 61028-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices; Measurement Techniques", June 2013
- IEC 62209-1, "Measurement procedure for the assessment of Specific Absorption Rate (SAR) from hand-held and body-mounted devices used next to the ear (frequency range of 300 MHz to 6 GHz)", July 2010
- IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 695664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor $k=2$, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D450V3-1103_Apr21 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed on leaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

Measurement Conditions
 DASY system configuration, as far as not given on page 1.

| | | |
|------------------------------|------------------------|-----------------------------|
| DASY Version | DASY5 | VIS: 10.4 |
| Extrapolation | Advanced Extrapolation | |
| Phantom | ELH Flat Phantom | Shell thickness: 2 ± 0.2 mm |
| Distance Dipole Center - TSL | 15 mm | with Spacer |
| Zoom Scan Resolution | dx, dy, dz = 5 mm | |
| Frequency | 450 MHz ± 1 MHz | |

Head TSL parameters
 The following parameters and calculations were applied.

| | Temperature | Permittivity | Conductivity |
|---|-----------------|--------------|------------------|
| Nominal Head TSL parameters | 22.0 °C | 43.5 | 0.87 nholm |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 43.1 ± 0.8 % | 0.87 nholm ± 0 % |
| Head TSL temperature change during test | < 0.5 °C | --- | --- |

SAR result with Head TSL

| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
|---|--------------------|--------------------------|
| SAR measured | 250 mW input power | 1.14 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 4.66 W/kg ± 18.1 % (k=2) |

| SAR averaged over 10 cm ³ (10 g) of Head TSL | Condition | |
|---|--------------------|--------------------------|
| SAR measured | 250 mW input power | 0.757 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 3.06 W/kg ± 17.6 % (k=2) |

Certificate No: D450V3-1103_Apr21 Page 2 of 6

DASY5 Validation Report for Head TSL

Test Laboratory: SPEAG, Zurich, Switzerland
 Date: 21.04.2021

DUT: Dipole 450 MHz; Type: D450V3; Serial: D450V3 - SN:1103

Communication System: UTD 0 - CW; Frequency: 450 MHz
 Medium parameters used: $f = 450 \text{ MHz}$; $\epsilon = 0.87 \text{ S/m}$; $\alpha = 43.1$; $\rho = 1000 \text{ kg/m}^3$
 Phantom section: Flat Section
 Measurement Standard: DASY5 (IEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: EX3DV4 - SN3877; ConvF(10.64, 10.64, 10.64) @ 450 MHz; Calibrated: 30.12.2020
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAF4 Sns54; Calibrated: 26.06.2020
- Phantom: ELI v4.0; Type: QDOVA001BB; Serial: TP.1003
- DASY52 S2.10.4(1S27); SEMCAD X 14.6.14(7483)

Dipole Calibration for Head Tissue/d=15mm, Pin=250mW/Zoom Scan (7x7x7)/Cube 0:
 Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 39.18 V/m; Power Drift = -0.08 dB
 Peak SAR (extrapolated) = 1.76 W/kg
 SAR(1g) = 1.14 W/kg; SAR(10g) = 0.767 W/kg
 Smallest distance from peaks to all points: 3 dB below: Larger than measurement grid
 Ratio of SAR at M2 to SAR at M1 = 54.9%
 Maximum value of SAR (measured) = 1.53 W/kg

0 dB = 1.53 W/kg = 1.85 dBW/kg

Certificate No: D450V3-1103_Apr21 Page 3 of 6

Appendix (Additional assessments outside the scope of SCS 0106)

Antenna Parameters with Head TSL

| | |
|--------------------------------------|-----------------|
| Impedance, transformed to feed point | 57.1 Ω - 2.6 jΩ |
| Return Loss | -23.6 dB |

General Antenna Parameters and Design

| | |
|----------------------------------|----------|
| Electrical Delay (one direction) | 1.346 ns |
|----------------------------------|----------|

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.
 The dipole is made of standard sevenkight coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.
 No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

Additional EUT Data

| | |
|-----------------|-------|
| Manufactured by | SPEAG |
|-----------------|-------|

Certificate No: D450V3-1103_Apr21 Page 4 of 6

Certificate No: D450V3-1103_Apr21 Page 6 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1.3 D750V3 - SN 1188

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4239613-2917 Fax: +86-10-4239613-2914
Email: cti@china.ttl.com.cn http://www.china.ttl.com.cn

Client: **SGS-CN** Certificate No: **Z22-60103**

CALIBRATION CERTIFICATE

Object: D750V3 - SN: 1188

Calibration Procedure(s): FF-Z11-003-01
Calibration Procedures for dipole validation kits

Calibration date: March 29, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22±2)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|------------------------|---------|--|-----------------------|
| Power Meter NRP2 | 108277 | 24-Sep-21 (CTTL No.J21X08320) | Sep-22 |
| Power sensor NRP8S | 104251 | 24-Sep-21 (CTTL No.J21X08320) | Sep-22 |
| Reference Probe EK30V4 | SN 7307 | 26-May-21 (SPEAG No.EK3-7307_May21) | May-22 |
| DAE4 | SN 1556 | 12-Jan-22 (CTTL-SPEAG No.Z22-60007) | Jan-23 |

| Secondary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|-------------------------|------------|--|-----------------------|
| Signal Generator E4439C | MY49571430 | 13-Jan-22 (CTTL No.J22X0409) | Jan-23 |
| Network Analyzer E5071C | MY48110873 | 14-Jan-22 (CTTL No.J22X0409) | Jan-23 |

Calibrated by: Zhao Jing, SAR Test Engineer (Signature)

Reviewed by: Lin Hao, SAR Test Engineer (Signature)

Approved by: Qi Dianyuan, SAR Project Leader (Signature)

Issued: April 3, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60103 Page 1 of 5

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4239613-2079 Fax: +86-10-4239613-2594
Email: cti@china.ttl.com.cn http://www.china.ttl.com.cn

Client: **SGS-CN** Certificate No: **Z22-60103**

CALIBRATION CERTIFICATE

Object: D750V3 - SN: 1188

Calibration Procedure(s): FF-Z11-003-01
Calibration Procedures for dipole validation kits

Calibration date: March 29, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22±2)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|------------------------|---------|--|-----------------------|
| Power Meter NRP2 | 108277 | 24-Sep-21 (CTTL No.J21X08320) | Sep-22 |
| Power sensor NRP8S | 104251 | 24-Sep-21 (CTTL No.J21X08320) | Sep-22 |
| Reference Probe EK30V4 | SN 7307 | 26-May-21 (SPEAG No.EK3-7307_May21) | May-22 |
| DAE4 | SN 1556 | 12-Jan-22 (CTTL-SPEAG No.Z22-60007) | Jan-23 |

| Secondary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|-------------------------|------------|--|-----------------------|
| Signal Generator E4439C | MY49571430 | 13-Jan-22 (CTTL No.J22X0409) | Jan-23 |
| Network Analyzer E5071C | MY48110873 | 14-Jan-22 (CTTL No.J22X0409) | Jan-23 |

Calibrated by: Zhao Jing, SAR Test Engineer (Signature)

Reviewed by: Lin Hao, SAR Test Engineer (Signature)

Approved by: Qi Dianyuan, SAR Project Leader (Signature)

Issued: April 3, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60103 Page 2 of 5

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4239613-2079 Fax: +86-10-4239613-2594
Email: cti@china.ttl.com.cn http://www.china.ttl.com.cn

Client: **SGS-CN** Certificate No: **Z22-60103**

CALIBRATION CERTIFICATE

Object: D750V3 - SN: 1188

Calibration Procedure(s): FF-Z11-003-01
Calibration Procedures for dipole validation kits

Calibration date: March 29, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22±2)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|------------------------|---------|--|-----------------------|
| Power Meter NRP2 | 108277 | 24-Sep-21 (CTTL No.J21X08320) | Sep-22 |
| Power sensor NRP8S | 104251 | 24-Sep-21 (CTTL No.J21X08320) | Sep-22 |
| Reference Probe EK30V4 | SN 7307 | 26-May-21 (SPEAG No.EK3-7307_May21) | May-22 |
| DAE4 | SN 1556 | 12-Jan-22 (CTTL-SPEAG No.Z22-60007) | Jan-23 |

| Secondary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|-------------------------|------------|--|-----------------------|
| Signal Generator E4439C | MY49571430 | 13-Jan-22 (CTTL No.J22X0409) | Jan-23 |
| Network Analyzer E5071C | MY48110873 | 14-Jan-22 (CTTL No.J22X0409) | Jan-23 |

Calibrated by: Zhao Jing, SAR Test Engineer (Signature)

Reviewed by: Lin Hao, SAR Test Engineer (Signature)

Approved by: Qi Dianyuan, SAR Project Leader (Signature)

Issued: April 3, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60103 Page 3 of 5

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4239613-2079 Fax: +86-10-4239613-2594
Email: cti@china.ttl.com.cn http://www.china.ttl.com.cn

Client: **SGS-CN** Certificate No: **Z22-60103**

CALIBRATION CERTIFICATE

Object: D750V3 - SN: 1188

Calibration Procedure(s): FF-Z11-003-01
Calibration Procedures for dipole validation kits

Calibration date: March 29, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22±2)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|------------------------|---------|--|-----------------------|
| Power Meter NRP2 | 108277 | 24-Sep-21 (CTTL No.J21X08320) | Sep-22 |
| Power sensor NRP8S | 104251 | 24-Sep-21 (CTTL No.J21X08320) | Sep-22 |
| Reference Probe EK30V4 | SN 7307 | 26-May-21 (SPEAG No.EK3-7307_May21) | May-22 |
| DAE4 | SN 1556 | 12-Jan-22 (CTTL-SPEAG No.Z22-60007) | Jan-23 |

| Secondary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|-------------------------|------------|--|-----------------------|
| Signal Generator E4439C | MY49571430 | 13-Jan-22 (CTTL No.J22X0409) | Jan-23 |
| Network Analyzer E5071C | MY48110873 | 14-Jan-22 (CTTL No.J22X0409) | Jan-23 |

Calibrated by: Zhao Jing, SAR Test Engineer (Signature)

Reviewed by: Lin Hao, SAR Test Engineer (Signature)

Approved by: Qi Dianyuan, SAR Project Leader (Signature)

Issued: April 3, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60103 Page 4 of 5



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

TTL Speaq Calibration Laboratory
 In Collaboration with **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@chinaul.com http://www.chinaul.com

DASY5 Validation Report for Head TSL Date: 2022-03-29
 Test Laboratory: CTTL, Beijing, China
 DUT: Dipole 750 MHz; Type: D750V3; Serial: D750V3-SG; 1188
 Communication System: UID 0; CW; Frequency: 750 MHz; Duty Cycle: 1:1
 Medium parameters used: $f = 750 \text{ MHz}$; $\epsilon = 0.888 \text{ S/m}$; $\alpha = 41.36$; $\rho = 1000 \text{ kg/m}^3$
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(10.31, 10.31) @ 750 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1556; Calibrated: 2022-01-12
- Phantom: MEP-V5-1C (20deg probe 03); Type: QD 000 P51 Cx; Serial: 1062
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7591)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 55.06 V/m; Power Drift = 0.00 dB
 Peak SAR (extrapolated) = 3.07 W/kg
 SAR(1 g) = 2.87 W/kg; SAR(10 g) = 1.37 W/kg
 Smallest distance from peaks to all points 3 dB below = 18.9 mm
 Ratio of SAR at M2 to SAR at M1 = 67.1%
 Maximum value of SAR (measured) = 2.74 W/kg

Certificate No: Z22-60103 Page 3 of 6

TTL Speaq Calibration Laboratory
 In Collaboration with **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@chinaul.com http://www.chinaul.com

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60103 Page 4 of 6

1.4 D835V2 - SN 4d114

TTL Speaq Calibration Laboratory
 In Collaboration with **CAICT**, **CNAN**, **CNAS**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@chinaul.com http://www.chinaul.com

Client: **SGS-CN** Certificate No: **Z22-60104**

CALIBRATION CERTIFICATE

Object: D835V2 - SN: 4d114
 Calibration Procedure(s): FF-Z11-003-01
 Calibration Procedures for dipole validation kits
 Calibration date: March 31, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility, environment temperature (22±3)°C and humidity <70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|------------------------|---------|--|-----------------------|
| Power Meter NRP2 | 100277 | 24-Sep-21 (CTTL No.J21X00326) | Sep-22 |
| Power sensor NRPBS | 104291 | 24-Sep-21 (CTTL No.J21X00326) | Sep-22 |
| Reference Probe EX3DV4 | SN 7307 | 26-May-21 (SPEAG No.EX3-7307_May21) | May-22 |
| DAE4 | SN 1506 | 12-Jan-22 (CTTL-SPEAG No.Z22-60007) | Jan-23 |

| Secondary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|-------------------------|------------|--|-----------------------|
| Signal Generator E4438C | MY46071430 | 13-Jan-22 (CTTL No.J22X00406) | Jan-23 |
| Network Analyzer E5071C | MY46110673 | 14-Jan-22 (CTTL No.J22X00406) | Jan-23 |

Calibrated by: Zhao Jing, SAR Test Engineer
 Reviewed by: Lin Hao, SAR Test Engineer
 Approved by: Qi Diaryuan, SAR Project Leader

Signature: [Signatures]
 Issued: April 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60104 Page 1 of 6

TTL Speaq Calibration Laboratory
 In Collaboration with **CAICT**

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
 E-mail: cti@chinaul.com http://www.chinaul.com

Glossary:
 TSL: tissue simulating liquid
 ConvF: sensitivity in TSL / NORMx,y,z
 N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:
 a) IEC/IEEE 62202-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
 b) KDB 685864, "SAR Measurement Requirements for 100 MHz to 5 GHz"

Additional Documentation:
 c) DASY4/S System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60104 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

In Collaboration with **TTL Speaq** CALIBRATION LABORATORY **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42204633-2079 Fax: +86-10-42204633-2504
 E-mail: csl@chinaeui.com http://www.chinaeui.com

Measurement Conditions
 DASYS system configuration, as far as not given on page 1.

| | | |
|------------------------------|--------------------------|-------------|
| DASYS Version | DASYS2 | V52.10.4 |
| Extrapolation | Advanced Extrapolation | |
| Phantom | Triple Flat Phantom 5.1C | |
| Distance Dipole Center - TSL | 15 mm | with Spacer |
| Zoom Scan Resolution | dx, dy, dz = 5 mm | |
| Frequency | 835 MHz ± 1 MHz | |

Head TSL parameters
 The following parameters and calculations were applied:

| | Temperature | Permittivity | Conductivity |
|---|-----------------|--------------|------------------|
| Nominal Head TSL parameters | 22.0 °C | 41.6 | 0.50 mho/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 41.0 ± 6 % | 0.91 mho/m ± 6 % |
| Head TSL temperature change during test | < 1.0 °C | --- | --- |

SAR result with Head TSL

| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
|---|--------------------|--------------------------|
| SAR measured | 250 mW input power | 2.37 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 6.40 W/kg ± 18.8 % (k=2) |
| SAR averaged over 10 cm ³ (10 g) of Head TSL | Condition | |
| SAR measured | 250 mW input power | 1.54 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 6.12 W/kg ± 16.7 % (k=2) |

Certificate No: 222-60104 Page 3 of 6

In Collaboration with **TTL Speaq** CALIBRATION LABORATORY **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42204633-2079 Fax: +86-10-42204633-2504
 E-mail: csl@chinaeui.com http://www.chinaeui.com

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL

| | |
|--------------------------------------|-----------------|
| Impedance, transformed to feed point | 48.70 - j22.0 Ω |
| Return Loss | -25.3dB |

General Antenna Parameters and Design

| | |
|----------------------------------|----------|
| Electrical Delay (one direction) | 1.307 ns |
|----------------------------------|----------|

After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is set according to the standards.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.

Additional EUT Data

| | |
|-----------------|-------|
| Manufactured by | SPEAG |
|-----------------|-------|

Certificate No: 222-60104 Page 4 of 6

In Collaboration with **TTL Speaq** CALIBRATION LABORATORY **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42204633-2079 Fax: +86-10-42204633-2504
 E-mail: csl@chinaeui.com http://www.chinaeui.com

DASYS Validation Report for Head TSL Date: 2022-03-31
 Test Laboratory: CTTL, Beijing, China
 DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 4d114
 Communication System: UID 0; CW; Frequency: 835 MHz; Duty Cycle: 1:1
 Medium parameters used: f = 835 MHz; σ = 0.007 S/m; ε_r = 40.98; ρ = 1000 kg/m³
 Phantom section: Right Section
 Measurement Standard: DASYS (IEEE/IEC/ANSI C63.19-2007)
 DASYS Configuration:

- Probe: EX3DV4 - SN7307; ConvF(10.13, 10.13, 10.13) @ 835 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA-E4 S01556; Calibrated: 2022-01-12
- Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASYS2 52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 57.88 V/m; Power Drift = 0.04 dB
 Peak SAR (extrapolated) = 3.56 W/kg
 SAR(1 g) = 2.37 W/kg; SAR(10 g) = 1.54 W/kg
 Smallest distance from peaks to all points 3 dB below = 15.8 mm
 Ratio of SAR at M2 to SAR at M1 = 66.2%
 Maximum value of SAR (measured) = 3.17 W/kg

0 dB = 3.17 W/kg = 5.01 dBW/kg

Certificate No: 222-60104 Page 5 of 6

In Collaboration with **TTL Speaq** CALIBRATION LABORATORY **CAICT**

Address: No. 52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42204633-2079 Fax: +86-10-42204633-2504
 E-mail: csl@chinaeui.com http://www.chinaeui.com

Impedance Measurement Plot for Head TSL

Certificate No: 222-60104 Page 6 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300 t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
 中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1.5 D900V2 - SN 1d079

| <p>Address: No.32 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42348633-3117 E-mail: csl@tts.com.cn</p> | <p>Address: No.32 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42348633-3117 E-mail: csl@tts.com.cn</p> | <p>Address: No.32 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42348633-3117 E-mail: csl@tts.com.cn</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|--|--|--|-----------------------|-------------------|--------|--|-----------------------|------------------------|----------|--------------------------------|--------------------------|----------------------------------|------------------------------|--------------------------------|-------------|------------------------|-------------------|--------------------------------------|-----------|-----------------|---------|--------------------------------------|-------------|---------------------|--------------|--|-----------------------|-------------------------|------------|--------------------------------|---------------|-------------------------|-------------------|---|---|-----|-----|---|-----------|-------|--------------|--------------------|-----------|-------------------------------------|------------------|--------------------------|---|-----------|-------|--------------|--------------------|-----------|-------------------------------------|------------------|--------------------------|
| <p>Client: SGS-CN Certificate No: Z22-40184</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>CALIBRATION CERTIFICATE</p> <p>Object: D900V2 - SN: 1d079</p> <p>Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits</p> <p>Calibration date: June 7, 2022</p> <p>The calibration Certificate documents the traceability to national standards, which realize the physical unity of measurements (5). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closed laboratory facility; environment temperature (23±1)°C and humidity <math>70\%</math>.</p> <p>Calibration Equipment used (M&TE orifice for calibration)</p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRPD</td> <td>106277</td> <td>24-Sep-21 (CTTL No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRPBS</td> <td>104291</td> <td>24-Sep-21 (CTTL No. J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EK3DVA</td> <td>SN 7464</td> <td>26-Jan-22 (SPEAG No. EK3-7464_Jan22)</td> <td>Jan-23</td> </tr> <tr> <td>DAEA</td> <td>SN 1656</td> <td>12-Jan-22 (CTTL SPEAG No. Z22-40007)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4438C</td> <td>M74907H430</td> <td>13-Jan-22 (CTTL No. J22X04096)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyser E5071C</td> <td>M749110673</td> <td>14-Jan-22 (CTTL No. J22X04096)</td> <td>Jan-23</td> </tr> </tbody> </table> <p>Calibrated by: Zhao Jing SAR Test Engineer</p> <p>Reviewed by: Lin Hao SAR Test Engineer</p> <p>Approved by: Qi Danyuan SAR Project Leader</p> <p>Issued: June 13, 2022</p> <p>The calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> | | | | Primary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration | Power Meter NRPD | 106277 | 24-Sep-21 (CTTL No. J21X08326) | Sep-22 | Power sensor NRPBS | 104291 | 24-Sep-21 (CTTL No. J21X08326) | Sep-22 | Reference Probe EK3DVA | SN 7464 | 26-Jan-22 (SPEAG No. EK3-7464_Jan22) | Jan-23 | DAEA | SN 1656 | 12-Jan-22 (CTTL SPEAG No. Z22-40007) | Jan-23 | Secondary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration | Signal Generator E4438C | M74907H430 | 13-Jan-22 (CTTL No. J22X04096) | Jan-23 | Network Analyser E5071C | M749110673 | 14-Jan-22 (CTTL No. J22X04096) | Jan-23 | | | | | | | | | | | | | | | | | | | | |
| Primary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power Meter NRPD | 106277 | 24-Sep-21 (CTTL No. J21X08326) | Sep-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor NRPBS | 104291 | 24-Sep-21 (CTTL No. J21X08326) | Sep-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Reference Probe EK3DVA | SN 7464 | 26-Jan-22 (SPEAG No. EK3-7464_Jan22) | Jan-23 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| DAEA | SN 1656 | 12-Jan-22 (CTTL SPEAG No. Z22-40007) | Jan-23 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Secondary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Signal Generator E4438C | M74907H430 | 13-Jan-22 (CTTL No. J22X04096) | Jan-23 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Network Analyser E5071C | M749110673 | 14-Jan-22 (CTTL No. J22X04096) | Jan-23 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>Certificate No: Z22-46184 Page 1 of 6</p> | | <p>Certificate No: Z22-46184 Page 2 of 6</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>Measurement Conditions</p> <p>DASY system configuration, as far as not given on page 1</p> <table border="1"> <thead> <tr> <th>DASY Version</th> <th>DASY52</th> <th>52.10.4</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>15 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>900 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> <p>Head TSL parameters</p> <p>The following parameters and calculations were applied:</p> <table border="1"> <thead> <tr> <th>Parameter</th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.9 °C</td> <td>41.0</td> <td>0.07 nS/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>22.0 ± 0.2 °C</td> <td>42.1 ± 0.5 %</td> <td>0.08 nS/m ± 0.6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><math>\pm 1.0 \text{ }^\circ\text{C}</math></td> <td>---</td> <td>---</td> </tr> </tbody> </table> <p>SAR result with Head TSL</p> <table border="1"> <thead> <tr> <th>SAR averaged over 1 cm³ (1 g) of Head TSL</th> <th>Condition</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>2.70 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>11.6 W/kg ± 19.8 % (k=2)</td> </tr> <tr> <td>SAR averaged over 18 cm³ (10 g) of Head TSL</td> <th>Condition</th> <th>Value</th> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>1.73 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>7.09 W/kg ± 19.7 % (k=2)</td> </tr> </tbody> </table> | | | | DASY Version | DASY52 | 52.10.4 | Extrapolation | Advanced Extrapolation | | Phantom | Triple Flat Phantom 5.1C | | Distance Dipole Center - TSL | 15 mm | with Spacer | Zoom Scan Resolution | dx, dy, dz = 5 mm | | Frequency | 900 MHz ± 1 MHz | | Parameter | Temperature | Permittivity | Conductivity | Nominal Head TSL parameters | 22.9 °C | 41.0 | 0.07 nS/m | Measured Head TSL parameters | 22.0 ± 0.2 °C | 42.1 ± 0.5 % | 0.08 nS/m ± 0.6 % | Head TSL temperature change during test | $\pm 1.0 \text{ }^\circ\text{C}$ | --- | --- | SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | Value | SAR measured | 250 mW input power | 2.70 W/kg | SAR for nominal Head TSL parameters | normalized to 1W | 11.6 W/kg ± 19.8 % (k=2) | SAR averaged over 18 cm ³ (10 g) of Head TSL | Condition | Value | SAR measured | 250 mW input power | 1.73 W/kg | SAR for nominal Head TSL parameters | normalized to 1W | 7.09 W/kg ± 19.7 % (k=2) |
| DASY Version | DASY52 | 52.10.4 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Extrapolation | Advanced Extrapolation | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Phantom | Triple Flat Phantom 5.1C | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Distance Dipole Center - TSL | 15 mm | with Spacer | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Zoom Scan Resolution | dx, dy, dz = 5 mm | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Frequency | 900 MHz ± 1 MHz | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Parameter | Temperature | Permittivity | Conductivity | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Nominal Head TSL parameters | 22.9 °C | 41.0 | 0.07 nS/m | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Measured Head TSL parameters | 22.0 ± 0.2 °C | 42.1 ± 0.5 % | 0.08 nS/m ± 0.6 % | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Head TSL temperature change during test | $\pm 1.0 \text{ }^\circ\text{C}$ | --- | --- | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | Value | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SAR measured | 250 mW input power | 2.70 W/kg | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SAR for nominal Head TSL parameters | normalized to 1W | 11.6 W/kg ± 19.8 % (k=2) | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SAR averaged over 18 cm ³ (10 g) of Head TSL | Condition | Value | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SAR measured | 250 mW input power | 1.73 W/kg | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SAR for nominal Head TSL parameters | normalized to 1W | 7.09 W/kg ± 19.7 % (k=2) | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>Certificate No: Z22-46184 Page 3 of 6</p> | | <p>Certificate No: Z22-40184 Page 4 of 6</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> <p>Antenna Parameters with Head TSL</p> <table border="1"> <thead> <tr> <th>Parameter</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Impedance, transformed to feed point</td> <td>48.10 - j6.40 Ω</td> </tr> <tr> <td>Return Loss</td> <td>-23.3 dB</td> </tr> </tbody> </table> <p>General Antenna Parameters and Design</p> <table border="1"> <thead> <tr> <th>Parameter</th> <th>Value</th> </tr> </thead> <tbody> <tr> <td>Electrical Delay (one-direction)</td> <td>1.312 ns</td> </tr> </tbody> </table> <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> <p>The dipole is made of standard semi-rigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> <p>Additional EUT Data</p> <table border="1"> <thead> <tr> <th>Manufacturer by</th> <th>SPEAG</th> </tr> </thead> <tbody> <tr> <td></td> <td></td> </tr> </tbody> </table> | | | | Parameter | Value | Impedance, transformed to feed point | 48.10 - j6.40 Ω | Return Loss | -23.3 dB | Parameter | Value | Electrical Delay (one-direction) | 1.312 ns | Manufacturer by | SPEAG | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Parameter | Value | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Impedance, transformed to feed point | 48.10 - j6.40 Ω | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Return Loss | -23.3 dB | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Parameter | Value | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Electrical Delay (one-direction) | 1.312 ns | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Manufacturer by | SPEAG | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No. 10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

TTL
CALIBRATION LABORATORY

Address: No.52 HuaYuanhui Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2187
E-mail: cti@ttda.net.cn

CAICT

中国认证
国际互认
校准
CNAS 13076

Date: 2022-06-07

DASY5 Validation Report for Head TSL

Test Laboratory: TTL, Beijing, China
DUT: Dipole 900 MHz; Type: D900V2; Serial: D900V2 - SN: 14079
 Communication System: UID 0, CW; Frequency: 900 MHz; Duty Cycle: 1:1
 Medium parameters used: $f = 900 \text{ MHz}$; $\sigma = 0.98 \text{ S/m}$; $\epsilon_r = 42.05$; $\rho = 1000 \text{ kg/m}^3$
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7464; Conv F(9.72, 9.72, 9.72) @ 900 MHz; Calibrated: 2022-01-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronic: DALE S81556; Calibrated: 2022-01-12
- Phantom: MFP_V5_1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52 S2.10.4(1335); SEMCAD X 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0; Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 59.81 V/m; Power Drift = -0.01 dB
 Peak SAR (extrapolated) = 4.20 W/kg
 SAR(1g) = 2.7% W/kg; SAR(10g) = 1.78 W/kg
 Smallest distance from peaks to all points, 3 dB below = 16 mm
 Ratio of SAR at M2 to SAR at M1 = 65.8%
 Maximum value of SAR (measured) = 3.71 W/kg

0 dB = 3.71 W/kg = 5.69 dBW/kg

Certificate No: Z22-60184 Page 3 of 6

TTL
CALIBRATION LABORATORY

Address: No.52 HuaYuanhui Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2187
E-mail: cti@ttda.net.cn

CAICT

中国认证
国际互认
校准
CNAS 13076

Date: 2022-06-07

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60184 Page 4 of 6

1.6 D1800V2 - SN 2d170

TTL
CALIBRATION LABORATORY

Address: No.52 HuaYuanhui Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2187
E-mail: cti@ttda.net.cn

CNAS

中国认证
国际互认
校准
CNAS 13076

CAICT

中国认证
国际互认
校准
CNAS 13076

Certificate No: Z22-60105

Client: SGS-CN

CALIBRATION CERTIFICATE

Object: **D1800V2 - SN: 2d170**

Calibration Procedure(s): **FF-Z11-003-01**
Calibration Procedures for dipole validation kits

Calibration date: **March 31, 2022**

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility, environment temperature (22±3)°C and humidity < 70%.

Calibration Equipment used (MATE critical for calibration)

| Primary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|------------------------|---------|--|-----------------------|
| Power Meter NRP2 | 106277 | 24-Sep-21 (CTTL No.J21X08326) | Sep-22 |
| Power sensor NRP5 | 104291 | 24-Sep-21 (CTTL No.J21X08326) | Sep-22 |
| Reference Probe EX3DV4 | SN 7307 | 26-May-21 (SPEAG No.EX3-7307_May21) | May-22 |
| DAEA | SN 1556 | 12-Jan-22 (CTTL-SPEAG No.Z22-60007) | Jan-23 |

| Secondary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|-------------------------|------------|--|-----------------------|
| Signal Generator E4438C | MY49071430 | 13-Jan-22 (CTTL No.Z22X0406) | Jan-23 |
| Network Analyzer E5071C | MY48110673 | 14-Jan-22 (CTTL No.Z22X0406) | Jan-23 |

Calibrated by: **Zhao Jing** (Name), SAR Test Engineer (Function), [Signature] (Signature)

Reviewed by: **Lin Hao** (Name), SAR Test Engineer (Function), [Signature] (Signature)

Approved by: **Qi Dianyan** (Name), SAR Project Leader (Function), [Signature] (Signature)

Issued April 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60105 Page 1 of 6

TTL
CALIBRATION LABORATORY

Address: No.52 HuaYuanhui Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2187
E-mail: cti@ttda.net.cn

CAICT

中国认证
国际互认
校准
CNAS 13076

Certificate No: Z22-60105

Glossary:

TSL: Issue simulating liquid
 ConvF: sensitivity in TSL / NORMx.y.z
 N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices, Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
- KDB 895664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

- DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60105 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300 t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
 中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

In Collaboration with
TTL Speag
 CALIBRATION LABORATORY

CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42394633-2079 Fax: +86-10-42394633-2504
 E-mail: cti@china.ttl.com.cn http://www.chinatst.com

Measurement Conditions
 DASYS system configuration, as far as not given on page 1

| | | |
|------------------------------|---------------------------|-------------|
| DASY Version | DASY52 | 52.10.4 |
| Extrapolation | Advanced Extrapolation | |
| Phantom | Triple Flat Phantoms 5.1C | |
| Distance Dipole Center - TSL | 10 mm | with Spacer |
| Zoom Scan Resolution | dx, dy, dz = 0.5 mm | |
| Frequency | 1800 MHz ± 1 MHz | |

Head TSL parameters
 The following parameters and calculations were applied:

| | | | |
|---|-----------------|--------------|--------------------|
| | Temperature | Permittivity | Conductivity |
| Nominal Head TSL parameters | 22.0 °C | 40.0 | 1.40 mho/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 40.8 ± 0.6 % | 1.41 mho/m ± 0.6 % |
| Head TSL temperature change during test | <1.0 °C | --- | --- |

SAR result with Head TSL

| | | |
|---|--------------------|--------------------------|
| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
| SAR measured | 250 mW input power | 9.73 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 38.9 W/kg ± 18.8 % (k=2) |
| SAR averaged over 10 cm ³ (10 g) of Head TSL | Condition | |
| SAR measured | 250 mW input power | 5.11 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 20.4 W/kg ± 18.7 % (k=2) |

Certificate No: Z22-60105 Page 3 of 6

In Collaboration with
TTL Speag
 CALIBRATION LABORATORY

CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42394633-2079 Fax: +86-10-42394633-2504
 E-mail: cti@china.ttl.com.cn http://www.chinatst.com

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL

| | |
|--------------------------------------|--------------|
| Impedance, transformed to feed point | 47.60-2.56jΩ |
| Return Loss | -29.4dB |

General Antenna Parameters and Design

| | |
|----------------------------------|----------|
| Electrical Delay (one direction) | 1.116 ns |
|----------------------------------|----------|

After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.

Additional EUT Data

| | |
|-----------------|-------|
| Manufactured by | SPEAG |
|-----------------|-------|

Certificate No: Z22-60105 Page 4 of 6

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60105 Page 6 of 6

In Collaboration with
TTL Speag
 CALIBRATION LABORATORY

CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-42394633-2079 Fax: +86-10-42394633-2504
 E-mail: cti@china.ttl.com.cn http://www.chinatst.com

DASY5 Validation Report for Head TSL Date: 2022-03-31
 Test Laboratory: CTTL, Beijing, China
 DUT: Dipole 1800 MHz; Type: D1800V2; Serial: D1800V2 - SN: Jd170
 Communication System: UID 0, CW; Frequency: 1800 MHz; Duty Cycle: 1:1
 Medium parameters used: f = 1800 MHz; σ = 1.411 S/m; ε = 40.62; ρ = 1000 kg/m³
 Phantom section: Right Section
 Measurement Standard: DASYS (IEEE/IEC/ANSI C63.19-2007)
 DASYS Configuration:

- Probe: EX3DV4 - SN7307; Coax(F)(R.34, R.34, R.34) @ 1800 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAIE4 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration-Zoom Scan (7x7x7) (7x7x7) Cube 0; Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 98.14 V/m; Power Drift = 0.03 dB
 Peak SAR (extrapolated) = 18.2 W/kg
 SAR(1 g) = 9.73 W/kg; SAR(10 g) = 5.11 W/kg
 Smallest distance from peaks to all points 3 dB below = 10 mm
 Ratio of SAR at M2 to SAR at M1 = 54%
 Maximum value of SAR (measured) = 15.2 W/kg

Certificate No: Z22-60105 Page 5 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days.
Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1.7 D1900V2 - SN 5d136

| <p>Address: No. 12 Huayuanli Road, Huailin District, Beijing, 100191, China Tel: +86-10-4236463-2117 E-mail: vtl@tts.com.cn</p> | <p>Address: No. 52 Huayuanli Road, Huailin District, Beijing, 100191, China Tel: +86-10-4236463-2117 E-mail: vtl@tts.com.cn</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|--|---|--|-----------------------|--|-----------------------|------------------|-----------|--------------------------------|--------|--------------------|---------|--------------------------------|--------|------------------------|------------|--------------------------------------|--------|------|---------|--------------------------------------|--------|---------------------|------|--|-----------------------|-------------------------|------------|--------------------------------|--------|-------------------------|------------|--------------------------------|--------|
| <p>Client: SGS-CN Certificate No: Z22-00185</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>CALIBRATION CERTIFICATE</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>Object: D1900V2 - SN 5d136</p> <p>Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits</p> <p>Calibration date: June 7, 2022</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>The calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>All calibrations have been conducted in the closed laboratory facility; environment temperature (23±1)°C and humidity 70%.</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>Calibration Equipment used (M&TE critical for calibration)</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2</td> <td>106277</td> <td>24-Sep-21 (CTTL No. Z21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRPES</td> <td>104291</td> <td>24-Sep-21 (CTTL No. Z21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EK3DV4</td> <td>SN 7684</td> <td>26-Jan-22 (SPEAG No. EK3-7684_Jan22)</td> <td>Jan-23</td> </tr> <tr> <td>DAE4</td> <td>SN 1656</td> <td>12-Jan-22 (CTTL SPEAG No. Z22-00007)</td> <td>Jan-23</td> </tr> </tbody> </table> <table border="1"> <thead> <tr> <th>Secondary Standards</th> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4438C</td> <td>MY48071430</td> <td>13-Jan-22 (CTTL No. Z22X00405)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyser E5071C</td> <td>MY48110673</td> <td>14-Jan-22 (CTTL No. Z22X00406)</td> <td>Jan-23</td> </tr> </tbody> </table> | | Primary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration | Power Meter NRP2 | 106277 | 24-Sep-21 (CTTL No. Z21X08326) | Sep-22 | Power sensor NRPES | 104291 | 24-Sep-21 (CTTL No. Z21X08326) | Sep-22 | Reference Probe EK3DV4 | SN 7684 | 26-Jan-22 (SPEAG No. EK3-7684_Jan22) | Jan-23 | DAE4 | SN 1656 | 12-Jan-22 (CTTL SPEAG No. Z22-00007) | Jan-23 | Secondary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration | Signal Generator E4438C | MY48071430 | 13-Jan-22 (CTTL No. Z22X00405) | Jan-23 | Network Analyser E5071C | MY48110673 | 14-Jan-22 (CTTL No. Z22X00406) | Jan-23 |
| Primary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power Meter NRP2 | 106277 | 24-Sep-21 (CTTL No. Z21X08326) | Sep-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor NRPES | 104291 | 24-Sep-21 (CTTL No. Z21X08326) | Sep-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Reference Probe EK3DV4 | SN 7684 | 26-Jan-22 (SPEAG No. EK3-7684_Jan22) | Jan-23 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| DAE4 | SN 1656 | 12-Jan-22 (CTTL SPEAG No. Z22-00007) | Jan-23 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Secondary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Signal Generator E4438C | MY48071430 | 13-Jan-22 (CTTL No. Z22X00405) | Jan-23 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Network Analyser E5071C | MY48110673 | 14-Jan-22 (CTTL No. Z22X00406) | Jan-23 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <table border="1"> <thead> <tr> <th>Calibrated by:</th> <th>Name</th> <th>Function</th> <th>Signature</th> </tr> </thead> <tbody> <tr> <td></td> <td>Zhao Jing</td> <td>SAR Test Engineer</td> <td></td> </tr> <tr> <td>Reviewed by:</td> <td>Lin Hao</td> <td>SAR Test Engineer</td> <td></td> </tr> <tr> <td>Approved by:</td> <td>Qi Dianyan</td> <td>SAR Project Leader</td> <td></td> </tr> </tbody> </table> <p>Issued June 13, 2022</p> | | Calibrated by: | Name | Function | Signature | | Zhao Jing | SAR Test Engineer | | Reviewed by: | Lin Hao | SAR Test Engineer | | Approved by: | Qi Dianyan | SAR Project Leader | | | | | | | | | | | | | | | | | |
| Calibrated by: | Name | Function | Signature | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| | Zhao Jing | SAR Test Engineer | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Reviewed by: | Lin Hao | SAR Test Engineer | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Approved by: | Qi Dianyan | SAR Project Leader | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| <p>Certificate No: Z22-00185 Page 1 of 4</p> | <p>Certificate No: Z22-00185 Page 1 of 4</p> | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |

| <p>Address: No. 12 Huayuanli Road, Huailin District, Beijing, 100191, China Tel: +86-10-4236463-2117 E-mail: vtl@tts.com.cn</p> | <p>Address: No. 52 Huayuanli Road, Huailin District, Beijing, 100191, China Tel: +86-10-4236463-2117 E-mail: vtl@tts.com.cn</p> | | | | | | | | | | | | | | | | | | |
|---|---|---|--------------------|--------------|---------------|------------------------|-------------|-------------------------------------|--------------------------|------------------------------|---|--------------|--------------------|---|--------------------|-------------|-------------------------------------|------------------|--------------------------|
| <p>Measurement Conditions</p> | | | | | | | | | | | | | | | | | | | |
| <p>DASY system configuration, as far as not given on page 1.</p> | | | | | | | | | | | | | | | | | | | |
| <table border="1"> <thead> <tr> <th>DASY Version</th> <th>DASY32</th> <th>52.10.4</th> </tr> </thead> <tbody> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom S.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>10 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>5x, 6y, 4z ± 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>1900 MHz ± 1 MHz</td> <td></td> </tr> </tbody> </table> | | DASY Version | DASY32 | 52.10.4 | Extrapolation | Advanced Extrapolation | | Phantom | Triple Flat Phantom S.1C | | Distance Dipole Center - TSL | 10 mm | with Spacer | Zoom Scan Resolution | 5x, 6y, 4z ± 5 mm | | Frequency | 1900 MHz ± 1 MHz | |
| DASY Version | DASY32 | 52.10.4 | | | | | | | | | | | | | | | | | |
| Extrapolation | Advanced Extrapolation | | | | | | | | | | | | | | | | | | |
| Phantom | Triple Flat Phantom S.1C | | | | | | | | | | | | | | | | | | |
| Distance Dipole Center - TSL | 10 mm | with Spacer | | | | | | | | | | | | | | | | | |
| Zoom Scan Resolution | 5x, 6y, 4z ± 5 mm | | | | | | | | | | | | | | | | | | |
| Frequency | 1900 MHz ± 1 MHz | | | | | | | | | | | | | | | | | | |
| <p>Head TSL parameters</p> | | | | | | | | | | | | | | | | | | | |
| <p>The following parameters and calculations were applied:</p> | | | | | | | | | | | | | | | | | | | |
| <table border="1"> <thead> <tr> <th>Nominal Head TSL parameters</th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td></td> <td>23.0 °C</td> <td>40.0</td> <td>1.40 nS/cm</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>22.0 ± 0.5 °C</td> <td>39.9 ± 0.6 %</td> <td>1.39 nS/cm ± 0.6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td>+1.0 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> | | Nominal Head TSL parameters | Temperature | Permittivity | Conductivity | | 23.0 °C | 40.0 | 1.40 nS/cm | Measured Head TSL parameters | 22.0 ± 0.5 °C | 39.9 ± 0.6 % | 1.39 nS/cm ± 0.6 % | Head TSL temperature change during test | +1.0 °C | --- | --- | | |
| Nominal Head TSL parameters | Temperature | Permittivity | Conductivity | | | | | | | | | | | | | | | | |
| | 23.0 °C | 40.0 | 1.40 nS/cm | | | | | | | | | | | | | | | | |
| Measured Head TSL parameters | 22.0 ± 0.5 °C | 39.9 ± 0.6 % | 1.39 nS/cm ± 0.6 % | | | | | | | | | | | | | | | | |
| Head TSL temperature change during test | +1.0 °C | --- | --- | | | | | | | | | | | | | | | | |
| <p>SAR result with Head TSL</p> | | | | | | | | | | | | | | | | | | | |
| <table border="1"> <thead> <tr> <th>SAR averaged over 1 cm² (1 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>9.65 (0.0g)</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>46.0 W/kg ± 18.8 % (k=2)</td> </tr> <tr> <td>SAR averaged over 10 cm² (10 g) of Head TSL</td> <td>Condition</td> <td></td> </tr> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>5.18 (0.0g)</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>26.8 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table> | | SAR averaged over 1 cm ² (1 g) of Head TSL | Condition | | SAR measured | 250 mW input power | 9.65 (0.0g) | SAR for nominal Head TSL parameters | normalized to 1W | 46.0 W/kg ± 18.8 % (k=2) | SAR averaged over 10 cm ² (10 g) of Head TSL | Condition | | SAR measured | 250 mW input power | 5.18 (0.0g) | SAR for nominal Head TSL parameters | normalized to 1W | 26.8 W/kg ± 18.7 % (k=2) |
| SAR averaged over 1 cm ² (1 g) of Head TSL | Condition | | | | | | | | | | | | | | | | | | |
| SAR measured | 250 mW input power | 9.65 (0.0g) | | | | | | | | | | | | | | | | | |
| SAR for nominal Head TSL parameters | normalized to 1W | 46.0 W/kg ± 18.8 % (k=2) | | | | | | | | | | | | | | | | | |
| SAR averaged over 10 cm ² (10 g) of Head TSL | Condition | | | | | | | | | | | | | | | | | | |
| SAR measured | 250 mW input power | 5.18 (0.0g) | | | | | | | | | | | | | | | | | |
| SAR for nominal Head TSL parameters | normalized to 1W | 26.8 W/kg ± 18.7 % (k=2) | | | | | | | | | | | | | | | | | |
| <p>Certificate No: Z22-00185 Page 3 of 4</p> | <p>Certificate No: Z22-00185 Page 3 of 4</p> | | | | | | | | | | | | | | | | | | |

| <p>Address: No. 12 Huayuanli Road, Huailin District, Beijing, 100191, China Tel: +86-10-4236463-2117 E-mail: vtl@tts.com.cn</p> | <p>Address: No. 52 Huayuanli Road, Huailin District, Beijing, 100191, China Tel: +86-10-4236463-2117 E-mail: vtl@tts.com.cn</p> | | | | |
|--|---|-------------------------------------|---------------|-------------|---------|
| <p>Appendix (Additional assessments outside the scope of CNAS L0570)</p> | | | | | |
| <p>Antenna Parameters with Head TSL</p> | | | | | |
| <table border="1"> <thead> <tr> <th>Impedance, transformed to feedpoint</th> <th>51.20 ± 7.56Ω</th> </tr> </thead> <tbody> <tr> <td>Return Loss</td> <td>-22.4dB</td> </tr> </tbody> </table> | | Impedance, transformed to feedpoint | 51.20 ± 7.56Ω | Return Loss | -22.4dB |
| Impedance, transformed to feedpoint | 51.20 ± 7.56Ω | | | | |
| Return Loss | -22.4dB | | | | |
| <p>General Antenna Parameters and Design</p> | | | | | |
| <table border="1"> <thead> <tr> <th>Electrical Delay (one direction)</th> <th>1.109 ns</th> </tr> </thead> </table> | | Electrical Delay (one direction) | 1.109 ns | | |
| Electrical Delay (one direction) | 1.109 ns | | | | |
| <p>After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.</p> | | | | | |
| <p>The dipole is made of standard omnigrid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as indicated in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.</p> | | | | | |
| <p>Additional EUT Data</p> | | | | | |
| <table border="1"> <thead> <tr> <th>Manufactured by</th> <th>SPEAG</th> </tr> </thead> </table> | | Manufactured by | SPEAG | | |
| Manufactured by | SPEAG | | | | |
| <p>Certificate No: Z22-00185 Page 4 of 4</p> | <p>Certificate No: Z22-00185 Page 4 of 4</p> | | | | |



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No. 10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

DASY5 Validation Report for Head TSL

Test Laboratory: TTL, Beijing, China
 DUT: Dipole 1900 MHz; Type: D1900V2; Serial: D1900V2 - SN: 10410
 Communication System: C/D 0, CW; Frequency: 1900 MHz; Duty Cycle: 1:1
 Medium parameters used: $f = 1900 \text{ MHz}$; $\epsilon = 1.385 \text{ S/m}$; $\sigma = 39.85$; $\rho = 1000 \text{ kg/m}^3$
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7464; ConF(R,18, 8,18, 8,18) @ 1900 MHz; Calibrated: 2022-01-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1556; Calibrated: 2022-01-12
- Phantom: MPP_V5_1C (20kg probe 0); Type: QD 000 P51 Cx; Serial: 1062
- DASY52_S2.10.4(1555); SEMCAD X 14.6.14(7561)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube (0; Measurement grid: dx=5mm, dy=5mm, dz=5mm)
 Reference Value = 99.99 V/m; Power Devit = 0.04 dB
 Peak SAR (extrapolated) = 18.6 W/kg
 SAR(1 g) = 9.95 W/kg; SAR(10 g) = 5.18 W/kg
 Smallest distance from peaks to all points 3 dB below = 9.2 mm
 Ratio of SAR at M2 to SAR at M1 = 44.1%
 Maximum value of SAR (measured) = 15.6 W/kg

0 dB = 15.6 W/kg = 11.93 dBW/kg

Impedance Measurement Plot for Head TSL

1.8 D2000V2 - SN 1041

Client: SGS-CN Certificate No: Z22-60186

CALIBRATION CERTIFICATE

Object: D2000V2 - SN: 1041
 Calibration Procedure(s): FF-Z11-003-01
 Calibration date: June 8, 2022

The calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (S). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility; environment temperature (23±1)°C and humidity < 70%.

Calibration Equipment Used (MATE critical for calibration)

| Primary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|------------------------|---------|--|-----------------------|
| Power Meter NRP9 | 106277 | 24-Sep-21 (CTL No. J21X06226) | Sep-22 |
| Power sensor NRP85 | 104291 | 24-Sep-21 (CTL No. J21X06226) | Sep-22 |
| Reference Probe EX3DV4 | SN 7464 | 25-Jan-22 (SPEAG No. EX3-7464-Jan22) | Jan-23 |
| DAE4 | SN 1656 | 12-Jan-22 (TTL-SPEAG No. Z22-60007) | Jan-23 |

| Secondary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|-------------------------|------------|--|-----------------------|
| Signal Generator E4438C | MY49071430 | 13-Jan-22 (CTL No. J22X00409) | Jan-23 |
| Network Analyzer E5071C | MY46110673 | 14-Jan-22 (CTL No. J22X00406) | Jan-23 |

Calibrated by: Zhao Jing (SAR Test Engineer)
 Reviewed by: Lin Hao (SAR Test Engineer)
 Approved by: Qi Dianyan (SAR Project Leader)

Issued: June 13, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Glossary:

TSL: Issue simulating liquid
 ConF: sensitivity in TSL / INORMx.yz
 N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:

- IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
- ICB 865964, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

c) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to the feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the normal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com

No. 10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

In Collaboration with
TTL
CALIBRATION LABORATORY

ADD: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4203683-2117
E-mail: csl@ttml.com.cn http://www.ttml.com.cn

CAICT

Measurement Conditions
DASY System configuration, see for details in page 1.

| | | |
|------------------------------|--------------------------|-------------|
| DASY Version | DASY32 | 52.10.4 |
| Extrapolation | Advanced Extrapolation | |
| Phantom | Triple Flat Phantom 5.1C | |
| Distance Dipole Center - TSL | 10 mm | with Spacer |
| Zoom Scan Resolution | dx, dy, dz = 5 mm | |
| Frequency | 2000 MHz ± 1 MHz | |

Head TSL parameters
The following parameters and calculations were applied:

| | Temperature | Permittivity | Conductivity |
|---|-----------------|--------------|-----------------|
| Nominal Head TSL parameters | 22.9 °C | 40.0 | 1.40 mS/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 40.2 ± 5 % | 1.30 mS/m ± 6 % |
| Head TSL temperature change during test | <1.0 °C | --- | --- |

SAR result with Head TSL

| SAR averaged over 1 cm ³ (g) of Head TSL | Condition | |
|---|--------------------|--------------------------|
| SAR measured | 250 mW input power | 10.4 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 41.8 W/kg ± 18.8 % (n=2) |
| SAR averaged over 10 cm ³ (10 g) of Head TSL | Condition | |
| SAR measured | 250 mW input power | 5.30 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 21.3 W/kg ± 18.7 % (n=2) |

Certificate No: Z22-40186 Page 1 of 6

In Collaboration with
TTL
CALIBRATION LABORATORY

ADD: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4203683-2117
E-mail: csl@ttml.com.cn http://www.ttml.com.cn

CAICT

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL

| | |
|--------------------------------------|---------------|
| Impedance, transformed to feed point | 48.42 ± 0.74Ω |
| Return Loss | -34.9dB |

General Antenna Parameters and Design

| | |
|----------------------------------|----------|
| Electrical Delay (one direction) | 1.088 ns |
|----------------------------------|----------|

After long term use with 100W radiated power, only a slight warming of the dipole near the feed point can be measured.

The dipole is made of standard semi-rigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed point may be damaged.

Additional EUT Data

| | |
|-----------------|-------|
| Manufactured by | SPEAG |
|-----------------|-------|

Certificate No: Z22-40186 Page 4 of 6

In Collaboration with
TTL
CALIBRATION LABORATORY

ADD: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4203683-2117
E-mail: csl@ttml.com.cn http://www.ttml.com.cn

CAICT

DASY5 Validation Report for Head TSL Date: 2022-06-06

Test Laboratory: CTTI, Beijing, China
DUT: Dipole 2000 MHz; Type: D2000V2; Serial: D2000V2 - SN: 1041
Communication System: LIID 0, CW; Frequency: 2000 MHz; Duty Cycle: 1:1
Median parameters used: f = 2000 MHz; α = 1.392 S/m; ε = 40.21; ρ = 1000 kg/m³
Phantom section: Right Section
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
DASY5 Configuration:

- Probe: EX3DV4 - SN7464; Conn F(R, L, R, L, R, L) @ 2000 MHz; Calibrated: 2022-01-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1556; Calibrated: 2022-01-12
- Phantom: MFP_V5.1C (20kg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52: 52.10.4(15.5); SEMCAD X 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0; Measurement grid: dx=5mm, dy=5mm, dz=5mm
Reference Value = 103.4 V/m; Power DUT = 0.03 dB
Peak SAR (extrapolated) = 19.6 W/kg
SAR(1g) = 10.4 W/kg; SAR(10g) = 5.3 W/kg
Smallest distance from peaks to all points 3 dB below = 9.1 mm
Ratio of SAR at M2 to SAR at M1 = 53.6%
Maximum value of SAR (measured) = 16.3 W/kg

0 dB = 16.3 W/kg = 12.12 dBW/kg

Certificate No: Z22-40186 Page 5 of 6

In Collaboration with
TTL
CALIBRATION LABORATORY

ADD: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-4203683-2117
E-mail: csl@ttml.com.cn http://www.ttml.com.cn

CAICT

Impedance Measurement Plot for Head TSL

Certificate No: Z22-40186 Page 6 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1.9 D2300V2 - SN 1096

| | | | |
|--|--|---|-----------------------|
| | | | |
| Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191 Tel: +86-10-42304633-2512 Fax: +86-10-42304633-2594 E-mail: cti@chinaul.com http://www.chinaul.com | | Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2594 E-mail: cti@chinaul.com http://www.chinaul.com | |
| Client: SGS-CN Certificate No: Z22-60106 | | | |
| CALIBRATION CERTIFICATE | | | |
| Object | D2300V2 - SN: 1096 | | |
| Calibration Procedure(s) | FF-Z11-003-01 Calibration Procedures for dipole validation kits | | |
| Calibration date: | March 31, 2022 | | |
| This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. | | | |
| All calibrations have been conducted in the closed laboratory facility: environment temperature (23±3)°C and humidity<70%. | | | |
| Calibration Equipment used (MSTE critical for calibration) | | | |
| Primary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
| Power Meter NRP2 | 108277 | 24-Sep-21 (CTTL No.J21X08326) | Sep-22 |
| Power sensor NRP8S | 104291 | 24-Sep-21 (CTTL No.J21X08326) | Sep-22 |
| Reference Probe EX3DV4 | SN 7307 | 26-May-21(SPEAG.No.EX3-7307_May21) | May-22 |
| DAE4 | SN 1556 | 12-Jan-22(CTTL-SPEAG.No.Z22-60007) | Jan-23 |
| Secondary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
| Signal Generator E4438C | MY49071430 | 13-Jan-22 (CTTL No.J22X00406) | Jan-23 |
| Network Analyzer E5071C | MY48110873 | 14-Jan-22 (CTTL No.J22X00406) | Jan-23 |
| Calibrated by: | Name | Function | Signature |
| | Zhao Jing | SAR Test Engineer | |
| Reviewed by: | Lin Hao | SAR Test Engineer | |
| Approved by: | Qi Dianyuan | SAR Project Leader | |
| Issued: April 6, 2022 | | | |
| This calibration certificate shall not be reproduced except in full without written approval of the laboratory. | | | |
| Certificate No: Z22-60106 | | Page 1 of 6 | |
| | | | |
| Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2594 E-mail: cti@chinaul.com http://www.chinaul.com | | Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2594 E-mail: cti@chinaul.com http://www.chinaul.com | |
| Measurement Conditions DASY system configuration, as far as not given on page 1 | | | |
| DASY Version | DASY52 | 52.10.4 | |
| Extrapolation | Advanced Extrapolation | | |
| Phantom | Triple Flat Phantom 5.1C | | |
| Distance Dipole Center - TSL | 10 mm | with Spacer | |
| Zoom Scan Resolution | dx, dy, dz = 5 mm | | |
| Frequency | 2300 MHz ± 1 MHz | | |
| Head TSL parameters The following parameters and calculations were applied. | | | |
| | Temperature | Permittivity | Conductivity |
| Nominal Head TSL parameters | 22.0 °C | 39.5 | 1.67 mho/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 39.8 ± 6 % | 1.70 mho/m ± 6 % |
| Head TSL temperature change during test | < 1.0 °C | — | — |
| SAR result with Head TSL | | | |
| SAR averaged over 1 cm ² (1 g) of Head TSL | Condition | | |
| SAR measured | 250 mW input power | 12.4 W/kg | |
| SAR for nominal Head TSL parameters | normalized to 1W | 49.2 W/kg ± 18.8 % (k=2) | |
| SAR averaged over 10 cm ² (10 g) of Head TSL | Condition | | |
| SAR measured | 250 mW input power | 5.88 W/kg | |
| SAR for nominal Head TSL parameters | normalized to 1W | 23.4 W/kg ± 18.7 % (k=2) | |
| Certificate No: Z22-60106 | | Page 1 of 6 | |
| | | | |
| Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2594 E-mail: cti@chinaul.com http://www.chinaul.com | | Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2594 E-mail: cti@chinaul.com http://www.chinaul.com | |
| Glossary: TSL Issue simulating liquid CorvF sensitivity in TSL / NORM _{x,y,z} N/A not applicable or not measured | | | |
| Calibration is performed According to the Following Standards: a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1:528 Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020 b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz" | | | |
| Additional Documentation: c) DASY4/5 System Handbook | | | |
| Methods Applied and Interpretation of Parameters: | | | |
| <ul style="list-style-type: none"> Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated. Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis. Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required. Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required. SAR measured: SAR measured at the stated antenna input power. SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector. SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result. | | | |
| The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%. | | | |
| Certificate No: Z22-60106 | | Page 2 of 6 | |
| | | | |
| Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2594 E-mail: cti@chinaul.com http://www.chinaul.com | | Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-3079 Fax: +86-10-42304633-2594 E-mail: cti@chinaul.com http://www.chinaul.com | |
| Appendix (Additional assessments outside the scope of CNAS L0570) | | | |
| Antenna Parameters with Head TSL | | | |
| Impedance, transformed to feed point | 49.20 - 4.56jΩ | | |
| Return Loss | -26.6dB | | |
| General Antenna Parameters and Design | | | |
| Electrical Delay (one direction) | 1.083 ns | | |
| After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured. | | | |
| The dipole is made of standard semi-rigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connectors near the feed-point may be damaged. | | | |
| Additional EUT Data | | | |
| Manufactured by | SPEAG | | |
| Certificate No: Z22-60106 | | Page 4 of 6 | |



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

TTL S P E A G CALIBRATION LABORATORY
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-6236633-3079 Fax: +86-10-6236633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Date: 2022-03-31

DASY5 Validation Report for Head TSL
 Test Laboratory: CTTL, Beijing, China
 DUT: Dipole 2300 MHz; Type: D2300V2; Serial: D2300V2 - SN: 1094
 Communication System: ULD 0; CW; Frequency: 2300 MHz; Duty Cycle: 1:1
 Phantom parameters used: $f = 2300$ MHz; $\sigma = 1.702$ S/m; $\epsilon = 39.77$; $\rho = 1000$ kg/m³
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(8.01, 8.01, 8.01) @ 2300 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA#4 S#1556; Calibrated: 2022-01-12
- Phantom: MFP_V5.1C (dipole probe kit); Type: QD 000 P51 Cx; Serial: 1062
- DASY52 52.10.4(1555); SEMCAD X 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 102.7 V/m; Power Drift = 0.00 dB
 Peak SAR (extrapolated) = 24.8 W/kg
 SAR(HI) g = 12.4 W/kg; SAR(LI) g = 5.88 W/kg
 Smallest distance from peaks to all points 3 dB below = 9 mm
 Ratio of SAR at M2 to SAR at M1 = 50.4%
 Maximum value of SAR (measured) = 20.3 W/kg

0 dB = 24.3 W/kg = 13.87 dBW/kg

Certificate No: Z22-60106 Page 5 of 6

TTL S P E A G CALIBRATION LABORATORY
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-6236633-3079 Fax: +86-10-6236633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60106 Page 6 of 6

1.10 D2450V2 - SN 817

TTL S P E A G CALIBRATION LABORATORY
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-6236633-3079 Fax: +86-10-6236633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Client: **SGS-CN** Certificate No: **Z22-60107**

CALIBRATION CERTIFICATE

Object: D2450V2 - SN: 817
 Calibration Procedure(s): FF-Z11-003-01
 Calibration date: April 1, 2022

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on this following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22±3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|-------------------------|---------|--|-----------------------|
| Power Meter: NRP2 | 108277 | 24-Sep-21 (CTTL No.J21X08320) | Sep-22 |
| Power sensor: NRP8S | 104291 | 24-Sep-21 (CTTL No.J21X08320) | Sep-22 |
| Reference Probe: EX3DV4 | SN 7307 | 26-May-21(SPEAG.No.EX3-7307_May21) | May-22 |
| DAE4 | SN 1556 | 12-Jan-22(CTTL-SPEAG.No.Z22-60007) | Jan-23 |

| Secondary Standards | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration |
|--------------------------|------------|--|-----------------------|
| Signal Generator: E4439C | MY4907430 | 13-Jan-22 (CTTL No. J22X00406) | Jan-23 |
| Network Analyzer: E5071C | MY48110873 | 14-Jan-22 (CTTL No. J22X00406) | Jan-23 |

Calibrated by: Zhao Jing, SAR Test Engineer
 Reviewed by: Lin Hao, SAR Test Engineer
 Approved by: Qi Dianyuan, SAR Project Leader

Issued: April 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60107 Page 1 of 6

TTL S P E A G CALIBRATION LABORATORY
CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-6236633-3079 Fax: +86-10-6236633-2504
 E-mail: cti@china.ttl.com http://www.china.ttl.com

Glossary:
 TSL: tissue simulating liquid
 ConvF: sensitivity in TSL / NORMx.yz
 N/A: not applicable or not measured

Calibration is Performed According to the Following Standards:
 a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020
 b) KDB 865864, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:
 c) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60107 Page 2 of 6



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留园学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

In Collaboration with **TTL Speag** CALIBRATION LABORATORY

CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
E-mail: cti@china.ttl.com http://www.china.ttl.com

Measurement Conditions
DASY system configuration, as far as not given on page 1

| | | |
|------------------------------|--------------------------|-------------|
| DASY Version | DASY52 | 52.10.4 |
| Extrapolation | Advanced Extrapolation | |
| Phantom | Triple Flat Phantom 5.1C | |
| Distance Dipole Center - TSL | 10 mm | with Spacer |
| Zoom Scan Resolution | dx, dy, dz = 5 mm | |
| Frequency | 2450 MHz ± 1 MHz | |

Head TSL parameters
The following parameters and calculations were applied:

| | Temperature | Permittivity | Conductivity |
|---|---------------|--------------|------------------|
| Nominal Head TSL parameters | 22.0 °C | 39.2 | 1.80 mho/m |
| Measured Head TSL parameters | 22.0 ± 0.2 °C | 39.5 ± 0 % | 1.79 mho/m ± 0 % |
| Head TSL temperature change during test | <±1.0 °C | --- | --- |

SAR result with Head TSL

| | | |
|---|--------------------|--------------------------|
| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
| SAR measured | 250 mW input power | 13.2 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 63.0 W/kg ± 18.8 % (k=2) |
| SAR averaged over 10 cm ³ (10 g) of Head TSL | Condition | |
| SAR measured | 250 mW input power | 6.15 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 24.7 W/kg ± 18.7 % (k=2) |

Certificate No: Z22-60107 Page 3 of 6

In Collaboration with **TTL Speag** CALIBRATION LABORATORY

CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
E-mail: cti@china.ttl.com http://www.china.ttl.com

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL

| | |
|--------------------------------------|--------------|
| Impedance, transformed to feed point | 52.10 ± 3.2Q |
| Return Loss | -28.5dB |

General Antenna Parameters and Design

| | |
|----------------------------------|----------|
| Electrical Delay (one direction) | 1.000 ns |
|----------------------------------|----------|

After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.

The dipole is made of standard serrigrid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.

Additional EUT Data

| | |
|-----------------|-------|
| Manufactured by | SPEAG |
|-----------------|-------|

Certificate No: Z22-60107 Page 4 of 6

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60107 Page 5 of 6

In Collaboration with **TTL Speag** CALIBRATION LABORATORY

CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
E-mail: cti@china.ttl.com http://www.china.ttl.com

DASY5 Validation Report for Head TSL Date: 2022-04-01

Test Laboratory: CTTL, Beijing, China

DUT: Dipole 2450 MHz; Type: D2450V2; Serial: D2450V2 - SN: 817

Communication System: UID 0, CW; Frequency: 2450 MHz; Duty Cycle: 1:1

Medium parameters used: f = 2450 MHz; $\epsilon = 1.79$ S/m; $\sigma = 39.52$; $\rho = 1000$ kg/m³

Phantom section: Right Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(7.75, 7.75, 7.75) @ 2450 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DA14 Set1556; Calibrated: 2022-01-12
- Phantom: MFP_V5.1C (20deg probe tilt); Type: QD 000 P51 Cx; Serial: 1062
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration Zoom Scan (7x7x7) (7x7x7) Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 104.6 V/m; Power Drift = -0.03 dB

Peak SAR (extrapolated) = 27.0 W/kg

SAR(1 g) = 13.2 W/kg; SAR(10 g) = 6.15 W/kg

Smallest distance from peaks to all points 3 dB below = 8.9 mm

Ratio of SAR at M2 to SAR at M1 = -0.2%

Maximum value of SAR (measured) = 22.1 W/kg

0 dB = 22.1 W/kg = 13.44 dBW/kg

Certificate No: Z22-60107 Page 1 of 6

In Collaboration with **TTL Speag** CALIBRATION LABORATORY

CAICT

Address: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China
Tel: +86-10-42304633-2079 Fax: +86-10-42304633-2504
E-mail: cti@china.ttl.com http://www.china.ttl.com

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60107 Page 6 of 6














Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

1.11 D2600V2 - SN 1158

|   | |   | | | | | | | | | | | | | | | | | | | |
|--|---|--|--|-----------------------|------------------------------------|-------------------------------|-----------|-------------------------------------|-------------------------------|------------------------------|--------------------------------|------------------------------------|------------------|---|------------------------------------|--------|-----------|------------------|--|--|--|
| Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191 Tel: +86-10-42304633-2912 Fax: +86-10-42304633-2904 E-mail: vt@china-test.com http://www.china-test.com | | Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2979 Fax: +86-10-42304633-2904 E-mail: vt@china-test.com http://www.china-test.com | | | | | | | | | | | | | | | | | | | |
| Client: SGS-CN Certificate No: Z22-60108 | | | | | | | | | | | | | | | | | | | | | |
| CALIBRATION CERTIFICATE | | | | | | | | | | | | | | | | | | | | | |
| Object: D2600V2 - SN: 1158 | | | | | | | | | | | | | | | | | | | | | |
| Calibration Procedure(s): FF-Z11-003-01 Calibration Procedures for dipole validation kits | | | | | | | | | | | | | | | | | | | | | |
| Calibration date: March 31, 2022 | | | | | | | | | | | | | | | | | | | | | |
| This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. | | | | | | | | | | | | | | | | | | | | | |
| All calibrations have been conducted in the closed laboratory facility: environment temperature (23±)°C and humidity<70%. | | | | | | | | | | | | | | | | | | | | | |
| Calibration Equipment used (MSTE critical for calibration) | | | | | | | | | | | | | | | | | | | | | |
| Primary Standards | <table border="1"> <thead> <tr> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power Meter NRP2 102377</td> <td>24-Sep-21 (CTTL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Power sensor NRP8S 104291</td> <td>24-Sep-21 (CTTL No.J21X08326)</td> <td>Sep-22</td> </tr> <tr> <td>Reference Probe EX3DVA SN 7307</td> <td>26-May-21(SPEAG.No.EX3-7307_May21)</td> <td>May-22</td> </tr> <tr> <td>DAE4 SN 1556</td> <td>12-Jan-22(CTTL-SPEAG.No.Z22-60007)</td> <td>Jan-23</td> </tr> </tbody> </table> | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration | Power Meter NRP2 102377 | 24-Sep-21 (CTTL No.J21X08326) | Sep-22 | Power sensor NRP8S 104291 | 24-Sep-21 (CTTL No.J21X08326) | Sep-22 | Reference Probe EX3DVA SN 7307 | 26-May-21(SPEAG.No.EX3-7307_May21) | May-22 | DAE4 SN 1556 | 12-Jan-22(CTTL-SPEAG.No.Z22-60007) | Jan-23 | | | | | |
| ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration | | | | | | | | | | | | | | | | | | | |
| Power Meter NRP2 102377 | 24-Sep-21 (CTTL No.J21X08326) | Sep-22 | | | | | | | | | | | | | | | | | | | |
| Power sensor NRP8S 104291 | 24-Sep-21 (CTTL No.J21X08326) | Sep-22 | | | | | | | | | | | | | | | | | | | |
| Reference Probe EX3DVA SN 7307 | 26-May-21(SPEAG.No.EX3-7307_May21) | May-22 | | | | | | | | | | | | | | | | | | | |
| DAE4 SN 1556 | 12-Jan-22(CTTL-SPEAG.No.Z22-60007) | Jan-23 | | | | | | | | | | | | | | | | | | | |
| Secondary Standards | <table border="1"> <thead> <tr> <th>ID #</th> <th>Cal Date (Calibrated by Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Signal Generator E4438C MY49071430</td> <td>13-Jan-22 (CTTL No.Z22X0409)</td> <td>Jan-23</td> </tr> <tr> <td>Network Analyzer E5071C MY46110673</td> <td>14-Jan-22 (CTTL No.Z22X0409)</td> <td>Jan-23</td> </tr> </tbody> </table> | ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration | Signal Generator E4438C MY49071430 | 13-Jan-22 (CTTL No.Z22X0409) | Jan-23 | Network Analyzer E5071C MY46110673 | 14-Jan-22 (CTTL No.Z22X0409) | Jan-23 | | | | | | | | | | | |
| ID # | Cal Date (Calibrated by Certificate No.) | Scheduled Calibration | | | | | | | | | | | | | | | | | | | |
| Signal Generator E4438C MY49071430 | 13-Jan-22 (CTTL No.Z22X0409) | Jan-23 | | | | | | | | | | | | | | | | | | | |
| Network Analyzer E5071C MY46110673 | 14-Jan-22 (CTTL No.Z22X0409) | Jan-23 | | | | | | | | | | | | | | | | | | | |
| Calibrated by: Zhao Jing SAR Test Engineer | | Signature:  | | | | | | | | | | | | | | | | | | | |
| Reviewed by: Lin Hao SAR Test Engineer | | Signature:  | | | | | | | | | | | | | | | | | | | |
| Approved by: Qi Diaryuan SAR Project Leader | | Signature:  | | | | | | | | | | | | | | | | | | | |
| Issued: April 6, 2022 This calibration certificate shall not be reproduced except in full without written approval of the laboratory. | | | | | | | | | | | | | | | | | | | | | |
| Certificate No: Z22-60108 Page 1 of 6 | | Certificate No: Z22-60108 Page 2 of 6 | | | | | | | | | | | | | | | | | | | |
|   | |   | | | | | | | | | | | | | | | | | | | |
| Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2979 Fax: +86-10-42304633-2904 E-mail: vt@china-test.com http://www.china-test.com | | Add: No.52 HuaYuanBei Road, Haidian District, Beijing, 100191, China Tel: +86-10-42304633-2979 Fax: +86-10-42304633-2904 E-mail: vt@china-test.com http://www.china-test.com | | | | | | | | | | | | | | | | | | | |
| Measurement Conditions DASY system configuration, as for as not given on page 1. | | | | | | | | | | | | | | | | | | | | | |
| <table border="1"> <tr> <td>DASY Version</td> <td>DASY52</td> <td>52.10.4</td> </tr> <tr> <td>Extrapolation</td> <td>Advanced Extrapolation</td> <td></td> </tr> <tr> <td>Phantom</td> <td>Triple Flat Phantom 5.1C</td> <td></td> </tr> <tr> <td>Distance Dipole Center - TSL</td> <td>10 mm</td> <td>with Spacer</td> </tr> <tr> <td>Zoom Scan Resolution</td> <td>dx, dy, dz = 5 mm</td> <td></td> </tr> <tr> <td>Frequency</td> <td>2600 MHz ± 1 MHz</td> <td></td> </tr> </table> | | DASY Version | DASY52 | 52.10.4 | Extrapolation | Advanced Extrapolation | | Phantom | Triple Flat Phantom 5.1C | | Distance Dipole Center - TSL | 10 mm | with Spacer | Zoom Scan Resolution | dx, dy, dz = 5 mm | | Frequency | 2600 MHz ± 1 MHz | | | |
| DASY Version | DASY52 | 52.10.4 | | | | | | | | | | | | | | | | | | | |
| Extrapolation | Advanced Extrapolation | | | | | | | | | | | | | | | | | | | | |
| Phantom | Triple Flat Phantom 5.1C | | | | | | | | | | | | | | | | | | | | |
| Distance Dipole Center - TSL | 10 mm | with Spacer | | | | | | | | | | | | | | | | | | | |
| Zoom Scan Resolution | dx, dy, dz = 5 mm | | | | | | | | | | | | | | | | | | | | |
| Frequency | 2600 MHz ± 1 MHz | | | | | | | | | | | | | | | | | | | | |
| Head TSL parameters The following parameters and calculations were applied. | | | | | | | | | | | | | | | | | | | | | |
| <table border="1"> <thead> <tr> <th></th> <th>Temperature</th> <th>Permittivity</th> <th>Conductivity</th> </tr> </thead> <tbody> <tr> <td>Nominal Head TSL parameters</td> <td>22.0 °C</td> <td>38.0</td> <td>1.56 mho/m</td> </tr> <tr> <td>Measured Head TSL parameters</td> <td>(22.0 ± 0.2) °C</td> <td>38.7 ± 0.9 %</td> <td>1.96 mho/m ± 6 %</td> </tr> <tr> <td>Head TSL temperature change during test</td> <td><1.0 °C</td> <td>---</td> <td>---</td> </tr> </tbody> </table> | | | Temperature | Permittivity | Conductivity | Nominal Head TSL parameters | 22.0 °C | 38.0 | 1.56 mho/m | Measured Head TSL parameters | (22.0 ± 0.2) °C | 38.7 ± 0.9 % | 1.96 mho/m ± 6 % | Head TSL temperature change during test | <1.0 °C | --- | --- | | | | |
| | Temperature | Permittivity | Conductivity | | | | | | | | | | | | | | | | | | |
| Nominal Head TSL parameters | 22.0 °C | 38.0 | 1.56 mho/m | | | | | | | | | | | | | | | | | | |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 38.7 ± 0.9 % | 1.96 mho/m ± 6 % | | | | | | | | | | | | | | | | | | |
| Head TSL temperature change during test | <1.0 °C | --- | --- | | | | | | | | | | | | | | | | | | |
| SAR result with Head TSL | | | | | | | | | | | | | | | | | | | | | |
| <table border="1"> <thead> <tr> <th>SAR averaged over 1 cm² (1 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>13.7 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>64.8 W/kg ± 18.8 % (k=2)</td> </tr> </tbody> </table> | | SAR averaged over 1 cm ² (1 g) of Head TSL | Condition | | SAR measured | 250 mW input power | 13.7 W/kg | SAR for nominal Head TSL parameters | normalized to 1W | 64.8 W/kg ± 18.8 % (k=2) | | | | | | | | | | | |
| SAR averaged over 1 cm ² (1 g) of Head TSL | Condition | | | | | | | | | | | | | | | | | | | | |
| SAR measured | 250 mW input power | 13.7 W/kg | | | | | | | | | | | | | | | | | | | |
| SAR for nominal Head TSL parameters | normalized to 1W | 64.8 W/kg ± 18.8 % (k=2) | | | | | | | | | | | | | | | | | | | |
| <table border="1"> <thead> <tr> <th>SAR averaged over 10 cm² (10 g) of Head TSL</th> <th>Condition</th> <th></th> </tr> </thead> <tbody> <tr> <td>SAR measured</td> <td>250 mW input power</td> <td>6.12 W/kg</td> </tr> <tr> <td>SAR for nominal Head TSL parameters</td> <td>normalized to 1W</td> <td>34.8 W/kg ± 18.7 % (k=2)</td> </tr> </tbody> </table> | | SAR averaged over 10 cm ² (10 g) of Head TSL | Condition | | SAR measured | 250 mW input power | 6.12 W/kg | SAR for nominal Head TSL parameters | normalized to 1W | 34.8 W/kg ± 18.7 % (k=2) | | | | | | | | | | | |
| SAR averaged over 10 cm ² (10 g) of Head TSL | Condition | | | | | | | | | | | | | | | | | | | | |
| SAR measured | 250 mW input power | 6.12 W/kg | | | | | | | | | | | | | | | | | | | |
| SAR for nominal Head TSL parameters | normalized to 1W | 34.8 W/kg ± 18.7 % (k=2) | | | | | | | | | | | | | | | | | | | |
| Appendix (Additional assessments outside the scope of CNAS L0670) | | | | | | | | | | | | | | | | | | | | | |
| Antenna Parameters with Head TSL | | | | | | | | | | | | | | | | | | | | | |
| <table border="1"> <tr> <td>Impedance, transformed to feed point</td> <td>49.50 - 6.49jΩ</td> </tr> <tr> <td>Return Loss</td> <td>-23.8dB</td> </tr> </table> | | Impedance, transformed to feed point | 49.50 - 6.49jΩ | Return Loss | -23.8dB | | | | | | | | | | | | | | | | |
| Impedance, transformed to feed point | 49.50 - 6.49jΩ | | | | | | | | | | | | | | | | | | | | |
| Return Loss | -23.8dB | | | | | | | | | | | | | | | | | | | | |
| General Antenna Parameters and Design | | | | | | | | | | | | | | | | | | | | | |
| <table border="1"> <tr> <td>Electrical Delay (one direction)</td> <td>1.053 ns</td> </tr> </table> | | Electrical Delay (one direction) | 1.053 ns | | | | | | | | | | | | | | | | | | |
| Electrical Delay (one direction) | 1.053 ns | | | | | | | | | | | | | | | | | | | | |
| After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured. | | | | | | | | | | | | | | | | | | | | | |
| The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard. No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged. | | | | | | | | | | | | | | | | | | | | | |
| Additional EUT Data | | | | | | | | | | | | | | | | | | | | | |
| <table border="1"> <tr> <td>Manufactured by</td> <td>SPEAG</td> </tr> </table> | | Manufactured by | SPEAG | | | | | | | | | | | | | | | | | | |
| Manufactured by | SPEAG | | | | | | | | | | | | | | | | | | | | |
| Certificate No: Z22-60108 Page 3 of 6 | | Certificate No: Z22-60108 Page 4 of 6 | | | | | | | | | | | | | | | | | | | |



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

TTL Speaq Calibration Laboratory
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanli Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62106633-2079 Fax: +86-10-62106633-2504
 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn

Date: 2022-03-31

DASY5 Validation Report for Head TSL
 Test Laboratory: CTTL, Beijing, China
 DUT: Dipole 2600 MHz; Type: D2600V2; Serial: D2600V2 - SN: 1158
 Communication System: UID 0, CW, Frequency: 2600 MHz, Duty Cycle: 1:1
 Medium parameters used: $f = 2600 \text{ MHz}$; $\sigma = 1.955 \text{ S/m}$; $\epsilon = 38.68$; $\rho = 1000 \text{ kg/m}^3$
 Phantom section: Right Section
 Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
 DASY5 Configuration:

- Probe: EX3DV4 - SN7307; ConvF(7.5, 7.5, 7.5) @ 2600 MHz; Calibrated: 2021-05-26
- Sensor-Surface: 1-Ann (Mechanical Surface Detection)
- Electronics: DA4 Sra1556; Calibrated: 2022-01-12
- Phantom: MFP_V5_IC (3dbeg probe kit); Type: QD 000 P51 Cx; Serial: 1062
- DASY52 52.104(1535); SEMCAD X 14.6.14(7501)

Dipole Calibration/Zoom Scan (7x7x7) (7x7x7) Cube 0; Measurement grid: dx=5mm, dy=5mm, dz=5mm
 Reference Value = 103.3 V/m; Power Drift = 0.04 dB
 Peak SAR (extrapolated) = 29.0 W/kg
 SAR(1 g) = 13.7 W/kg; SAR(10 g) = 6.12 W/kg
 Smallest distance from peaks to all points 3 dB below = 8.9 mm
 Ratio of SAR at M2 to SAR at M1 = 47.5%
 Maximum value of SAR (measured) = 23.4 W/kg

Certificate No: Z22-60108 Page 1 of 6

TTL Speaq Calibration Laboratory
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanli Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62106633-2079 Fax: +86-10-62106633-2504
 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn

Impedance Measurement Plot for Head TSL

Certificate No: Z22-60108 Page 4 of 6

1.12 D5GHzV2 - SN 1095

TTL Speaq Calibration Laboratory
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanli Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62106633-2079 Fax: +86-10-62106633-2504
 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn

Client: **SGS-CN** Certificate No: **Z22-60187**

CALIBRATION CERTIFICATE

Object: **D5GHzV2 - SN 1095**

Calibration Procedure(s): **FF-211-003-01**
 Calibration Procedures for dipole validation kits

Calibration date: **June 1, 2022**

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility; environment temperature (22±3)°C and humidity < 70%.

Calibration Equipment used (MSTE critical for calibration)

| Primary Standards | ID # | Cal Date (Calibrated by: Certificate No.) | Scheduled Calibration |
|------------------------|---------|---|-----------------------|
| Power Meter NR92 | 106277 | 24-Sep-21 (CTTL No. J21X008326) | Sep-22 |
| Power sensor NR91S | 104291 | 24-Sep-21 (CTTL No. J21X008326) | Sep-22 |
| Reference Probe EX3DV4 | SN 7464 | 28-Jan-22(SPEAG No. EX3-7464_Jan22) | Jan-23 |
| DAE4 | SN 1656 | 12-Jan-22(CTTL-SPEAG No. Z22-60007) | Jan-23 |

| Secondary Standards | ID # | Cal Date (Calibrated by: Certificate No.) | Scheduled Calibration |
|-------------------------|------------|---|-----------------------|
| Signal Generator E4330C | MY49071430 | 13-Jan-22 (CTTL No. J22X00406) | Jan-23 |
| Network Analyzer E5071C | MY46110873 | 14-Jan-22 (CTTL No. J22X00406) | Jan-23 |

Calibrated by: **Zhao Jing** SAR Test Engineer

Reviewed by: **Lin Hao** SAR Test Engineer

Approved by: **Qi Dianyan** SAR Project Leader

Issued: June 6, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z22-60187 Page 1 of 10

TTL Speaq Calibration Laboratory
 In Collaboration with **CAICT**

Address: No. 52 HuaYuanli Road, Haidian District, Beijing, 100191, China
 Tel: +86-10-62106633-2079 Fax: +86-10-62106633-2504
 E-mail: cti@china.ttl.com.cn http://www.china.ttl.com.cn

Glossary:

TSL Issue simulating liquid sensitivity in TSL / NCRM_{k,y,z}

Conf# N/A not applicable or not measured

Calibration is Performed According to the Following Standards:

a) IEC/IEEE 62209-1528, "Measurement Procedure for The Assessment of Specific Absorption Rate of Human Exposure to Radio Frequency Fields from Hand-held and Body-mounted Wireless Communication Devices- Part 1528: Human Models, Instrumentation and Procedures (Frequency range of 4 MHz to 10 GHz)", October 2020

b) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Additional Documentation:

c) DASY4/5 System Handbook

Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: Z22-60187 Page 2 of 10



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com

No. 10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留园学生创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

In Collaboration with
TTL
CALIBRATION LABORATORY

ADD: No.52 Huayuanli Road, Haidian District, Beijing, 100191, China
Tel: +86-10-43202117
E-mail: cti@ttml.com.cn http://www.ttml.com.cn

CAICT

Measurement Conditions
SASV system configuration, as far as not given on page 1.

| | | |
|------------------------------|--|----------------------------------|
| DASY Version | DASY2 | 52.10.4 |
| Extrapolation | Advanced Extrapolation | |
| Phantoms | Triple Flat Phantoms 5.1C | |
| Distance Dipole Center - TSL | 10 mm | with Spacer |
| Zoom Scan Resolution | dx, dy = 4 mm, dz = 1.4 mm | Graded Ratio = 1.4 (Z-direction) |
| Frequency | 5200 MHz ± 1 MHz 5300 MHz ± 1 MHz 5500 MHz ± 1 MHz 5600 MHz ± 1 MHz | |

Head TSL parameters at 5200MHz
The following parameters and calculations were applied:

| | | | |
|---|-----------------|--------------|------------------|
| Nominal Head TSL parameters | Temperature | Permittivity | Conductivity |
| | 22.0 °C | 35.0 | 4.66 mho/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 35.4 ± 6 % | 4.62 mho/m ± 6 % |
| Head TSL temperature change during test | <+1.0 °C | --- | --- |

SAR result with Head TSL at 5200MHz

| | | |
|---|--------------------|--------------------------|
| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
| SAR measured | 250 mW input power | 7.79 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 77.6 W/kg ± 24.4 % (k=2) |
| SAR averaged over 10 cm ³ (10 g) of Head TSL | Condition | |
| SAR measured | 250 mW input power | 2.22 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 22.1 W/kg ± 24.2 % (k=2) |

Certificate No: Z22-60187 Page 3 of 10

In Collaboration with
TTL
CALIBRATION LABORATORY

ADD: No.52 Huayuanli Road, Haidian District, Beijing, 100191, China
Tel: +86-10-43202117
E-mail: cti@ttml.com.cn http://www.ttml.com.cn

CAICT

Head TSL parameters at 5300MHz
The following parameters and calculations were applied:

| | | | |
|---|-----------------|--------------|------------------|
| Nominal Head TSL parameters | Temperature | Permittivity | Conductivity |
| | 22.0 °C | 35.9 | 4.76 mho/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 35.2 ± 6 % | 4.73 mho/m ± 6 % |
| Head TSL temperature change during test | <+1.0 °C | --- | --- |

SAR result with Head TSL at 5300MHz

| | | |
|---|--------------------|--------------------------|
| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
| SAR measured | 100 mW input power | 7.94 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 75.1 W/kg ± 24.4 % (k=2) |
| SAR averaged over 10 cm ³ (10 g) of Head TSL | Condition | |
| SAR measured | 100 mW input power | 2.27 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 22.6 W/kg ± 24.2 % (k=2) |

Head TSL parameters at 5500MHz
The following parameters and calculations were applied:

| | | | |
|---|-----------------|--------------|------------------|
| Nominal Head TSL parameters | Temperature | Permittivity | Conductivity |
| | 22.0 °C | 35.6 | 4.96 mho/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 34.9 ± 6 % | 4.94 mho/m ± 6 % |
| Head TSL temperature change during test | <+1.0 °C | --- | --- |

SAR result with Head TSL at 5500MHz

| | | |
|---|--------------------|--------------------------|
| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
| SAR measured | 100 mW input power | 8.29 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 82.9 W/kg ± 24.4 % (k=2) |
| SAR averaged over 10 cm ³ (10 g) of Head TSL | Condition | |
| SAR measured | 100 mW input power | 2.34 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 23.3 W/kg ± 24.2 % (k=2) |

Certificate No: Z22-60187 Page 4 of 10

In Collaboration with
TTL
CALIBRATION LABORATORY

ADD: No.52 Huayuanli Road, Haidian District, Beijing, 100191, China
Tel: +86-10-43202117
E-mail: cti@ttml.com.cn http://www.ttml.com.cn

CAICT

Head TSL parameters at 5600MHz
The following parameters and calculations were applied:

| | | | |
|---|-----------------|--------------|------------------|
| Nominal Head TSL parameters | Temperature | Permittivity | Conductivity |
| | 22.0 °C | 35.5 | 5.07 mho/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 34.7 ± 6 % | 5.05 mho/m ± 6 % |
| Head TSL temperature change during test | <+1.0 °C | --- | --- |

SAR result with Head TSL at 5600MHz

| | | |
|---|--------------------|--------------------------|
| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
| SAR measured | 100 mW input power | 8.12 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 80.8 W/kg ± 24.4 % (k=2) |
| SAR averaged over 10 cm ³ (10 g) of Head TSL | Condition | |
| SAR measured | 100 mW input power | 2.30 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 22.9 W/kg ± 24.2 % (k=2) |

Head TSL parameters at 5800MHz
The following parameters and calculations were applied:

| | | | |
|---|-----------------|--------------|------------------|
| Nominal Head TSL parameters | Temperature | Permittivity | Conductivity |
| | 22.0 °C | 35.3 | 5.27 mho/m |
| Measured Head TSL parameters | (22.0 ± 0.2) °C | 34.4 ± 6 % | 5.25 mho/m ± 6 % |
| Head TSL temperature change during test | <+1.0 °C | --- | --- |

SAR result with Head TSL at 5800MHz

| | | |
|---|--------------------|--------------------------|
| SAR averaged over 1 cm ³ (1 g) of Head TSL | Condition | |
| SAR measured | 100 mW input power | 7.71 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 76.7 W/kg ± 24.4 % (k=2) |
| SAR averaged over 10 cm ³ (10 g) of Head TSL | Condition | |
| SAR measured | 100 mW input power | 2.16 W/kg |
| SAR for nominal Head TSL parameters | normalized to 1W | 21.6 W/kg ± 24.2 % (k=2) |

Certificate No: Z22-60187 Page 5 of 10

In Collaboration with
TTL
CALIBRATION LABORATORY

ADD: No.52 Huayuanli Road, Haidian District, Beijing, 100191, China
Tel: +86-10-43202117
E-mail: cti@ttml.com.cn http://www.ttml.com.cn

CAICT

Appendix (Additional assessments outside the scope of CNAS L0570)

Antenna Parameters with Head TSL at 5200MHz

| | |
|--------------------------------------|----------------|
| Impedance, transformed to feed point | 46.10 - 5.03jΩ |
| Return Loss | -23.6dB |

Antenna Parameters with Head TSL at 5300MHz

| | |
|--------------------------------------|----------------|
| Impedance, transformed to feed point | 47.60 - 2.42jΩ |
| Return Loss | -29.5dB |

Antenna Parameters with Head TSL at 5500MHz

| | |
|--------------------------------------|----------------|
| Impedance, transformed to feed point | 50.30 - 4.36jΩ |
| Return Loss | -27.4dB |

Antenna Parameters with Head TSL at 5600MHz

| | |
|--------------------------------------|----------------|
| Impedance, transformed to feed point | 54.50 - 4.89jΩ |
| Return Loss | -24.0dB |

Antenna Parameters with Head TSL at 5800MHz

| | |
|--------------------------------------|----------------|
| Impedance, transformed to feed point | 51.50 - 5.61jΩ |
| Return Loss | -24.9dB |

Certificate No: Z22-60187 Page 6 of 10



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiyue Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

In Collaboration with
TTL
CALIBRATION LABORATORY

Addr: No.52 HuaYuanRoad, Haidian District, Beijing, 100191, China
Tel: +86-10-42020117
E-mail: csl@ttslab.com

General Antenna Parameters and Design

| | |
|----------------------------------|----------|
| Electrical Delay (one direction) | 1.101 ns |
|----------------------------------|----------|

After long term use with 100W radiated power, only a slight warming of the dipole near the feed-point can be measured.

The dipole is made of standard semi-rigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.
No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feed-point may be damaged.

Additional EUT Data

| | |
|-----------------|-------|
| Manufactured by | SPEAG |
|-----------------|-------|

Certificate No: Z22-60187 Page 1 of 10

In Collaboration with
TTL
CALIBRATION LABORATORY

Addr: No.52 HuaYuanRoad, Haidian District, Beijing, 100191, China
Tel: +86-10-42020117
E-mail: csl@ttslab.com

DASY5 Validation Report for Head TSL Date: 2022-04-01

Test Laboratory: CTTL, Beijing, China
DUT: Dipole 5GHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1095

Communication System: CW; Frequency: 5200 MHz; Frequency: 5300 MHz; Frequency: 5500 MHz; Frequency: 5600 MHz; Frequency: 5800 MHz; Duty Cycle: 1:1
Medium parameters used: f = 5200 MHz; $\sigma = 4.62$ S/m; $\epsilon_r = 35.39$; $\rho = 1000$ kg/m³
Medium parameters used: f = 5300 MHz; $\sigma = 4.73$ S/m; $\epsilon_r = 35.19$; $\rho = 1000$ kg/m³
Medium parameters used: f = 5500 MHz; $\sigma = 4.939$ S/m; $\epsilon_r = 34.83$; $\rho = 1000$ kg/m³
Medium parameters used: f = 5600 MHz; $\sigma = 5.051$ S/m; $\epsilon_r = 34.69$; $\rho = 1000$ kg/m³
Medium parameters used: f = 5800 MHz; $\sigma = 5.247$ S/m; $\epsilon_r = 34.42$; $\rho = 1000$ kg/m³
Phantom section: Right Section
Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)
DASY5 Configuration:

- Probe: EX3DV4 - SN7464; ConvF(5.6, 5.6, 5.6) @ 5200 MHz; ConvF(5.32, 5.32, 5.32) @ 5300 MHz; ConvF(5.11, 5.11, 5.11) @ 5500 MHz; ConvF(4.91, 4.91, 4.91) @ 5600 MHz; ConvF(5, 5, 5) @ 5800 MHz; Calibrated: 2022-01-26
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn1556; Calibrated: 2022-01-12
- Phantom: MPF_V5.1C (20deg probe TR); Type: GD 000 P51 Cx; Serial: 1062
- DASY52 52.10.4(1535); SEMCAD X 14.6.14(7601)

Dipole Calibration /Pin=100mW, d=10mm, f=5200 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 60.60 V/m; Power Drift = -0.06 dB
Peak SAR (extrapolated) = 29.8 W/kg
SAR(1 g) = 7.75 W/kg; SAR(10 g) = 2.22 W/kg
Smallest distance from peaks to all points 3 dB below = 7.2 mm
Ratio of SAR at M2 to SAR at M1 = 66.8%
Maximum value of SAR (measured) = 18.3 W/kg

Dipole Calibration /Pin=100mW, d=10mm, f=5300 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 61.08 V/m; Power Drift = -0.07 dB
Peak SAR (extrapolated) = 31.5 W/kg
SAR(1 g) = 7.94 W/kg; SAR(10 g) = 2.27 W/kg
Smallest distance from peaks to all points 3 dB below = 7.2 mm
Ratio of SAR at M2 to SAR at M1 = 65.5%
Maximum value of SAR (measured) = 19.0 W/kg

Certificate No: Z22-60187 Page 4 of 10


In Collaboration with
TTL
CALIBRATION LABORATORY

Addr: No.52 HuaYuanRoad, Haidian District, Beijing, 100191, China
Tel: +86-10-42020117
E-mail: csl@ttslab.com

Dipole Calibration /Pin=100mW, d=10mm, f=5500 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 61.92 V/m; Power Drift = -0.08 dB
Peak SAR (extrapolated) = 34.7 W/kg
SAR(1 g) = 8.29 W/kg; SAR(10 g) = 2.34 W/kg
Smallest distance from peaks to all points 3 dB below = 7.2 mm
Ratio of SAR at M2 to SAR at M1 = 63.5%
Maximum value of SAR (measured) = 20.2 W/kg

Dipole Calibration /Pin=100mW, d=10mm, f=5600 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 65.08 V/m; Power Drift = -0.07 dB
Peak SAR (extrapolated) = 35.2 W/kg
SAR(1 g) = 8.12 W/kg; SAR(10 g) = 2.3 W/kg
Smallest distance from peaks to all points 3 dB below = 7.2 mm
Ratio of SAR at M2 to SAR at M1 = 62.5%
Maximum value of SAR (measured) = 19.1 W/kg

Dipole Calibration /Pin=100mW, d=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0; Measurement grid: dx=4mm, dy=4mm, dz=1.4mm
Reference Value = 62.13 V/m; Power Drift = -0.06 dB
Peak SAR (extrapolated) = 34.8 W/kg
SAR(1 g) = 7.71 W/kg; SAR(10 g) = 2.16 W/kg
Smallest distance from peaks to all points 3 dB below = 7.2 mm
Ratio of SAR at M2 to SAR at M1 = 61.6%
Maximum value of SAR (measured) = 18.7 W/kg




Certificate No: Z22-60187 Page 9 of 10

In Collaboration with
TTL
CALIBRATION LABORATORY

Addr: No.52 HuaYuanRoad, Haidian District, Beijing, 100191, China
Tel: +86-10-42020117
E-mail: csl@ttslab.com

Impedance Measurement Plot for Head TSL



Certificate No: Z22-60187 Page 10 of 10



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgs.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

2 DAE4 - SN 1245

S P E A G

IMPORTANT NOTICE

USAGE OF THE DAE4

The DAE unit is a delicate, high precision instrument and requires careful treatment by the user. There are no serviceable parts inside the DAE. Special attention shall be given to the following points:

Battery Exchange: The battery cover of the DAE4 unit is fixed using a screw, over tightening the screw may cause the threads inside the DAE to wear out.

Shipping of the DAE: Before shipping the DAE to SPEAG for calibration, remove the batteries and pack the DAE in an anti-static bag. This anti-static bag shall then be packed into a larger box or container which protects the DAE from impacts during transportation. The package shall be marked to indicate that a fragile instrument is inside.

E-Stop Failures: Touch detection may be malfunctioning due to broken magnets in the E-stop. Rough handling of the E-stop may lead to damage of these magnets. Touch and collision errors are often caused by dust and dirt accumulated in the E-stop. To prevent E-stop failure, the customer shall always mount the probe to the DAE carefully and keep the DAE unit in a non-dusty environment if not used for measurements.

Repair: Minor repairs are performed at no extra cost during the annual calibration. However, SPEAG reserves the right to charge for any repair especially if rough unprofessional handling caused the defect.

DASY Configuration Files: Since the exact values of the DAE input resistances, as measured during the calibration procedure of a DAE unit, are not used by the DASY software, a nominal value of 200 MΩ/cm is given in the corresponding configuration file.

Important Note:
Warranty and calibration is void if the DAE unit is disassembled partly or fully by the Customer.

Important Note:
Never attempt to grease or oil the E-stop assembly. Cleaning and readjusting of the E-stop assembly is allowed by certified SPEAG personnel only and is part of the annual calibration procedure.

Important Note:
To prevent damage of the DAE probe connector pins, use great care when installing the probe to the DAE. Carefully connect the probe with the connector notch oriented in the mating position. Avoid any rotational movement of the probe body versus the DAE while turning the locking nut of the connector. The same care shall be used when disconnecting the probe from the DAE.

TN_EH190306AE_DAE4.docx 07.03.2019

Calibration Laboratory of
Schmid & Partner
Engineering AG
Zugzwilstrasse 43, 8004 Zurich, Switzerland

Schweizerischer Kalibrierdienst
Service suisse d'étalonnage
Servizio svizzero di taratura
Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 0108

Client: **SGS-CN (Auden)** Certificate No.: **DAE4-1245_May22**

CALIBRATION CERTIFICATE

Object: **DAE4 - SD 000 D04 BM - SN: 1245**

Calibration procedure(s): **QA CAL-06 v30**
Calibration procedure for the data acquisition electronics (DAE)

Calibration date: **May 30, 2022**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 0.1 °C and humidity < 70%.

Calibration Equipment used (MPE critical for calibration)

| Primary Standards | ID # | Cal Date (Certificate No.) | Scheduled Calibration |
|-------------------------------|-------------|----------------------------|-----------------------|
| Kalibray Multimeter Type 2001 | SN: 0810279 | 31-Aug-21 (No:31306) | Aug-22 |

| Secondary Standards | ID # | Check Date (in house) | Scheduled Check |
|---------------------------|--------------------|----------------------------|-----------------------|
| Auto DAE Calibration Unit | SE LWS 231 AA 1001 | 24-Jan-22 (in house check) | In house check Jan-23 |
| Calibrator Blue V2 1 | SE LMS 006 AA 1002 | 24-Jan-22 (in house check) | In house check Jan-23 |

Calibrated by: **Dominique Stufen** (Name), **Laboratory Technician** (Function), *[Signature]* (Signature)

Approved by: **Bruno Klein** (Name), **Technical Manager** (Function), *[Signature]* (Signature)

Issued: May 30, 2022

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: DAE4-1245_May22 Page 1 of 5

Calibration Laboratory of
Schmid & Partner
Engineering AG
Zugzwilstrasse 43, 8004 Zurich, Switzerland

Schweizerischer Kalibrierdienst
Service suisse d'étalonnage
Servizio svizzero di taratura
Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)
The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: SCS 0108

Glossary

DAE: data acquisition electronics
Connector angle: information used in DASY system to align probe sensor X to the robot coordinate system.

Methods Applied and Interpretation of Parameters

- DC Voltage Measurement:** Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- Connector angle:** The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The following parameters as documented in the Appendix contain technical information as a result from the performance test and require no uncertainty.
 - DC Voltage Measurement Linearity:** Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement.
 - Common mode sensitivity:** Influence of a positive or negative common mode voltage on the differential measurement.
 - Channel separation:** Influence of a voltage on the neighbor channels not subject to an input voltage.
 - AD Converter Values with inputs shorted:** Values on the internal AD converter corresponding to zero input voltage
 - Input Offset Measurement:** Output voltage and statistical results over a large number of zero voltage measurements.
 - Input Offset Current:** Typical value for information; Maximum channel input offset current, not considering the input resistance.
 - Input resistance:** Typical value for information; DAE input resistance at the connector, during internal auto-zeroing and during measurement.
 - Low Battery Alarm Voltage:** Typical value for information. Below this voltage, a battery alarm signal is generated.
 - Power consumption:** Typical value for information. Supply currents in various operating modes.

Certificate No: DAE4-1245_May22 Page 2 of 5

DC Voltage Measurement

AD - Converter Resolution nominal
High Range: 1LSB = 6.1 μV, Full range = -100...+500 mV
Low Range: 1LSB = 61 μV, Full range = -1...+50 mV
DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

| Calibration Factors | X | Y | Z |
|---------------------|-----------------------|-----------------------|-----------------------|
| High Range | 405.265 ± 0.02% (k=2) | 403.974 ± 0.02% (k=2) | 406.062 ± 0.02% (k=2) |
| Low Range | 3.96534 ± 1.50% (k=2) | 3.99568 ± 1.50% (k=2) | 4.01015 ± 1.50% (k=2) |

Connector Angle

| Connector Angle to be used in DASY system | 30.0° ± 1° |
|---|------------|
|---|------------|

Certificate No: DAE4-1245_May22 Page 3 of 5



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300 t(86-512)57355888 f(86-512)57370818 www.sgsgr.com.cn
 中国·江苏·昆山市留园伟业路10号 邮编 215300 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

| Appendix (Additional assessments outside the scope of SCS0108) | | | | |
|---|--------------------------------|---------------------------------|--------------------------------|---------------------|
| 1. DC Voltage Linearity | | | | |
| DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec | | | | |
| High Range | Reading (µV) | Difference (µV) | Error (%) | |
| Channel X + Input | 10964.45 | 1.62 | 0.20 | |
| Channel X - Input | 2004.56 | 2.22 | -0.91 | |
| Channel Y + Input | -2005.14 | 1.12 | -0.61 | |
| Channel Y - Input | 19964.72 | 1.68 | 0.00 | |
| Channel Z + Input | 2001.22 | -1.00 | -0.50 | |
| Channel Z - Input | -2003.25 | -1.27 | -0.21 | |
| Channel X + Input | 10962.84 | 0.19 | 0.00 | |
| Channel X - Input | 2003.09 | 0.88 | 0.60 | |
| Channel Y + Input | -2001.73 | -0.27 | 0.00 | |
| Channel Y - Input | 2011.61 | 0.41 | 0.22 | |
| Channel X + Input | 202.54 | 0.65 | 0.32 | |
| Channel X - Input | -197.86 | 0.67 | -0.34 | |
| Channel Y + Input | 2002.05 | 0.58 | 0.03 | |
| Channel Y - Input | 201.27 | -0.67 | -0.28 | |
| Channel Z + Input | -196.23 | -0.05 | 0.03 | |
| Channel Z - Input | 2007.36 | 0.08 | 0.00 | |
| Channel X + Input | 200.03 | -1.63 | -0.78 | |
| Channel X - Input | -196.85 | -1.57 | -0.79 | |
| 2. Common mode sensitivity | | | | |
| DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec | | | | |
| | Common mode input Voltage (µV) | High Range Average Reading (µV) | Low Range Average Reading (µV) | |
| Channel X | 200 | -5.87 | -7.69 | |
| | -200 | 9.12 | 7.75 | |
| Channel Y | 200 | -8.68 | -8.28 | |
| | -200 | 8.82 | 6.36 | |
| Channel Z | 200 | -5.36 | -5.60 | |
| | -200 | 3.58 | 3.96 | |
| 3. Channel separation | | | | |
| DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec | | | | |
| | Input Voltage (µV) | Channel X (µV) | Channel Y (µV) | Channel Z (µV) |
| Channel X | 200 | - | 4.07 | -3.14 |
| Channel Y | 200 | 9.36 | - | 4.27 |
| Channel Z | 200 | 10.11 | 7.14 | - |
| Certificate No: DA64-1245_May22 Page 4 of 5 | | | | |
| 4. AD-Converter Values with inputs shorted | | | | |
| DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec | | | | |
| | High Range (LSB) | Low Range (LSB) | | |
| Channel X | 10984 | 17010 | | |
| Channel Y | 16562 | 15788 | | |
| Channel Z | 16023 | 15658 | | |
| 5. Input Offset Measurement | | | | |
| DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec | | | | |
| | Average (µV) | min. Offset (µV) | max. Offset (µV) | Std. Deviation (µV) |
| Channel X | 1.00 | -0.15 | 1.93 | 0.45 |
| Channel Y | -0.18 | -1.28 | 0.94 | 0.45 |
| Channel Z | -0.58 | -0.81 | 0.58 | 0.60 |
| 6. Input Offset Current | | | | |
| Nominal input unity offset current on all channels: $-25\mu A$ | | | | |
| 7. Input Resistance (Typical values for information) | | | | |
| | Zeroing (MΩin) | Measuring (MΩin) | | |
| Channel X | 200 | 200 | | |
| Channel Y | 200 | 200 | | |
| Channel Z | 200 | 200 | | |
| 8. Low Battery Alarm Voltage (Typical values for information) | | | | |
| Typical values: Alarm Level (VDC) | | | | |
| Supply (+ Vcc) | | | +7.9 | |
| Supply (- Vcc) | | | -7.8 | |
| 9. Power Consumption (Typical values for information) | | | | |
| Typical values: Switched off (mA), Stand by (mA), Transmitting (mA) | | | | |
| Supply (+ Vcc) | -0.01 | +6 | +14 | |
| Supply (- Vcc) | -0.01 | -8 | -8 | |
| Certificate No: DA64-1245_May22 Page 5 of 5 | | | | |

3 EX3DV4 - SN 7346

| <p>Calibration Laboratory of Schmid & Partner Engineering AG Zugstrasse 41, 8964 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Client: Auden Certificate No: EX3-7346_Mar22</p> <p>CALIBRATION CERTIFICATE</p> <p>Client: EX3DV4 - SN 7346</p> <p>Calibration procedure(s): QA CAL-01-V9, QA CAL-14-V6, QA CAL-23-V5, QA CAL-25-V7 Calibration procedure for dosimetric E-field probes</p> <p>Calibration date: March 30, 2022</p> <p>This calibration certificate documents the traceability to national standards, which require the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.</p> <p>All calibrations have been conducted in the closest laboratory facility, environment temperature 22 ± 0.1°C and humidity <math>< 10\%</math></p> <p>Calibration Equipment used (MATE critical for calibration)</p> <table border="1"> <thead> <tr> <th>Primary Standards</th> <th>ID</th> <th>Cal Date (Certificate No.)</th> <th>Scheduled Calibration</th> </tr> </thead> <tbody> <tr> <td>Power meter 160P</td> <td>SN: 134719</td> <td>09-Apr-21 (No. 211-0201-103050)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor MP2-201</td> <td>SN: 103214</td> <td>09-Apr-21 (No. 211-0201-103211)</td> <td>Apr-22</td> </tr> <tr> <td>Power sensor MP2-201</td> <td>SN: 103243</td> <td>09-Apr-21 (No. 211-0201-103242)</td> <td>Apr-22</td> </tr> <tr> <td>Reference 20.06 reference</td> <td>SN: 000780 (201)</td> <td>09-Apr-21 (No. 211-0201-103243)</td> <td>Apr-22</td> </tr> <tr> <td>SATA</td> <td>SN: 880</td> <td>13-Oct-21 (No. 216-0460-04601)</td> <td>Oct-22</td> </tr> <tr> <td>Reference Probe E302</td> <td>SN: 3013</td> <td>27-Oct-21 (No. 216-0461-04611)</td> <td>Oct-22</td> </tr> </tbody> </table> <p>Reference Standards</p> <table border="1"> <thead> <tr> <th>ID</th> <th>Check Date (in house)</th> <th>Scheduled Check</th> </tr> </thead> <tbody> <tr> <td>Power meter E4138</td> <td>09-Apr-21 (in house check Jun-23)</td> <td>in house check Jun-23</td> </tr> <tr> <td>Power sensor E4138A</td> <td>SN: M14-149887</td> <td>09-Apr-21 (in house check Jun-23)</td> </tr> <tr> <td>Power sensor E4138A</td> <td>SN: 00011215</td> <td>09-Apr-21 (in house check Jun-23)</td> </tr> <tr> <td>SI generator 40-8646C</td> <td>SN: 020402011740</td> <td>09-Apr-21 (in house check Jun-23)</td> </tr> <tr> <td>Network Analyzer 83056A</td> <td>SN: U841030417</td> <td>31-Mar-14 (in house check Oct-23)</td> </tr> </tbody> </table> <p>Calibrated by: Ben Kuhn (Lab Manager) Signature: <i>[Signature]</i></p> <p>Approved by: Ben Kuhn (Deputy Manager) Signature: <i>[Signature]</i></p> <p>Issued: March 31, 2022</p> <p>This calibration certificate shall not be reproduced except in full without written approval of the laboratory.</p> <p>Certificate No: EX3-7346_Mar22 Page 1 of 24</p> | Primary Standards | ID | Cal Date (Certificate No.) | Scheduled Calibration | Power meter 160P | SN: 134719 | 09-Apr-21 (No. 211-0201-103050) | Apr-22 | Power sensor MP2-201 | SN: 103214 | 09-Apr-21 (No. 211-0201-103211) | Apr-22 | Power sensor MP2-201 | SN: 103243 | 09-Apr-21 (No. 211-0201-103242) | Apr-22 | Reference 20.06 reference | SN: 000780 (201) | 09-Apr-21 (No. 211-0201-103243) | Apr-22 | SATA | SN: 880 | 13-Oct-21 (No. 216-0460-04601) | Oct-22 | Reference Probe E302 | SN: 3013 | 27-Oct-21 (No. 216-0461-04611) | Oct-22 | ID | Check Date (in house) | Scheduled Check | Power meter E4138 | 09-Apr-21 (in house check Jun-23) | in house check Jun-23 | Power sensor E4138A | SN: M14-149887 | 09-Apr-21 (in house check Jun-23) | Power sensor E4138A | SN: 00011215 | 09-Apr-21 (in house check Jun-23) | SI generator 40-8646C | SN: 020402011740 | 09-Apr-21 (in house check Jun-23) | Network Analyzer 83056A | SN: U841030417 | 31-Mar-14 (in house check Oct-23) | <p>Calibration Laboratory of Schmid & Partner Engineering AG Zugstrasse 41, 8964 Zurich, Switzerland</p> <p>Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates</p> <p>Client: Auden Certificate No: EX3-7346_Mar22</p> <p>Glossary:</p> <p>TSL: tissue simulating liquid NORM_{M,y,z}: sensitivity in free space Conf: crest factor (1 duty cycle) of the RF signal modulation dependent linearization parameters A, B, C, D: rotation around probe axis Polarization: rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e. $\theta = 0$ is normal to probe axis Connector Angle: information used in DASY system to align probe sensor X to the robot coordinate system</p> <p>Calibration is Performed According to the Following Standards:</p> <p>a) IEC60729-1038: "Measurement Procedure For the Assessment Of Specific Absorption Rate Of Human Exposure To Radio Frequency Fields From Hand-Held And Body-Worn Wireless Communication Devices - Part 1038: Human Models, Implementation And Procedures (Frequency Range of 4 MHz to 10 GHz)"; October 2002</p> <p>b) IEC608666: "SAR Measurement Requirements for 100 MHz to 6 GHz"</p> <p>Methods Applied and Interpretation of Parameters:</p> <ul style="list-style-type: none"> NORM_{M,y,z}: Assessed for E-field polarization $\theta = 0$ ($f = 900$ MHz in TEM-cell; $f = 1800$ MHz: R22 waveguide) NORM_{M,y,z} are only intermediate values, i.e., the uncertainties of NORM_{M,y,z} does not affect the E-field uncertainty model TSL (see below Conf). NORM_{M,y,z} = NORM_{M,y,z} * frequency response (see Frequency Response Chart). This linearization is implemented in DASY software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of Conf. DCP: DCP are non-linear linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media. PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics. Any z: Any z: C₁, z₁; D₁, z₁; V₁, z₁; A, B, C, D are non-linear linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. V₁ is the maximum calibration range expressed in RMS voltage across the probe. Conf: Conf Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for $f < 600$ MHz) and trade waveguide using analytical field distributions based on power measurements to $f = 600$ MHz. The same setups are used for assessments of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY for GAD_{M,y,z} and GAD_{M,y,z} * Conf_{M,y,z} whereas the uncertainty corresponds to that given for Conf. A frequency dependent Conf is used in DASY version 4.4 and higher which allows extending the validity from 1.50 MHz to 1.100 MHz. Spherical latracy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna. Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required. Connector Angle: The angle is assessed using the information gained by determining the NORM_M (no uncertainty required). <p>Certificate No: EX3-7346_Mar22 Page 2 of 24</p> |
|---|-----------------------------------|-----------------------------------|----------------------------|-----------------------|------------------|------------|---------------------------------|--------|----------------------|------------|---------------------------------|--------|----------------------|------------|---------------------------------|--------|---------------------------|------------------|---------------------------------|--------|------|---------|--------------------------------|--------|----------------------|----------|--------------------------------|--------|----|-----------------------|-----------------|-------------------|-----------------------------------|-----------------------|---------------------|----------------|-----------------------------------|---------------------|--------------|-----------------------------------|-----------------------|------------------|-----------------------------------|-------------------------|----------------|-----------------------------------|---|
| Primary Standards | ID | Cal Date (Certificate No.) | Scheduled Calibration | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power meter 160P | SN: 134719 | 09-Apr-21 (No. 211-0201-103050) | Apr-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor MP2-201 | SN: 103214 | 09-Apr-21 (No. 211-0201-103211) | Apr-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor MP2-201 | SN: 103243 | 09-Apr-21 (No. 211-0201-103242) | Apr-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Reference 20.06 reference | SN: 000780 (201) | 09-Apr-21 (No. 211-0201-103243) | Apr-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SATA | SN: 880 | 13-Oct-21 (No. 216-0460-04601) | Oct-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Reference Probe E302 | SN: 3013 | 27-Oct-21 (No. 216-0461-04611) | Oct-22 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| ID | Check Date (in house) | Scheduled Check | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power meter E4138 | 09-Apr-21 (in house check Jun-23) | in house check Jun-23 | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor E4138A | SN: M14-149887 | 09-Apr-21 (in house check Jun-23) | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Power sensor E4138A | SN: 00011215 | 09-Apr-21 (in house check Jun-23) | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| SI generator 40-8646C | SN: 020402011740 | 09-Apr-21 (in house check Jun-23) | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Network Analyzer 83056A | SN: U841030417 | 31-Mar-14 (in house check Oct-23) | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com

EX3DV4 - SN:7346 March 30, 2022

DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346

Basic Calibration Parameters

| Norm. $\mu V/V/mV^{1/2}$ | Norm. $\mu V/mV^{1/2}$ | Norm. $\mu V/mV^{1/2}$ | Norm. $\mu V/mV^{1/2}$ |
|--------------------------|------------------------|------------------------|------------------------|
| 0.46 | 0.47 | 0.61 | $\pm 10.1\%$ |
| 191.4 | 196.0 | 106.9 | |

Calibration Results for Modulation Response

| UID | Communication System Name | Modulation | Modulation Rate | Modulation Depth | Modulation Index | Modulation Error | Modulation Error (%) |
|-------|-----------------------------|------------|-----------------|------------------|------------------|------------------|-------------------------------|
| 0 | QW | X | 0.00 | 0.00 | 1.00 | 0.00 | 143.5 $\pm 3.0\%$ |
| 1003A | Pulse Waveform (200Hz, 10%) | X | 3.33 | 66.90 | 17.46 | 10.00 | 80.0 $\pm 3.5\%$ $\pm 0.8\%$ |
| 1003A | AAA | Y | 4.63 | 79.70 | 19.35 | 96.0 | |
| 1003A | AAA | Z | 1.83 | 61.25 | 6.75 | 96.0 | |
| 1003A | AAA | X | 3.00 | 70.83 | 11.33 | 639 | 80.0 $\pm 2.4\%$ $\pm 0.8\%$ |
| 1003A | AAA | Y | 11.51 | 81.32 | 14.72 | 80.0 | |
| 1003A | AAA | Z | 0.93 | 60.00 | 3.17 | 80.0 | |
| 1003A | Pulse Waveform (200Hz, 40%) | X | 7.41 | 79.85 | 12.51 | 3.94 | 80.0 $\pm 2.7\%$ $\pm 0.8\%$ |
| 1003A | AAA | Y | 20.09 | 81.42 | 16.51 | 96.0 | |
| 1003A | AAA | Z | 0.18 | 138.38 | 3.01 | 96.0 | |
| 1003A | Pulse Waveform (200Hz, 60%) | X | 3.97 | 75.12 | 9.52 | 2.12 | 120.0 $\pm 1.7\%$ $\pm 0.8\%$ |
| 1003A | AAA | Y | 20.09 | 91.38 | 16.29 | 120.0 | |
| 1003A | AAA | Z | 1.94 | 139.51 | 16.87 | 120.0 | |
| 1003A | GPSK Waveform, 1 MHz | X | 1.47 | 64.88 | 13.42 | 1.00 | 150.0 $\pm 4.2\%$ $\pm 0.8\%$ |
| 1003A | AAA | Y | 3.66 | 66.37 | 14.65 | 1.00 | |
| 1003A | AAA | Z | 0.45 | 61.88 | 11.05 | 1.00 | |
| 1003A | GPSK Waveform, 10 MHz | X | 1.36 | 66.37 | 14.65 | 0.00 | 150.0 $\pm 1.1\%$ $\pm 0.8\%$ |
| 1003A | AAA | Y | 2.06 | 67.33 | 15.58 | 0.00 | |
| 1003A | AAA | Z | 0.21 | 64.25 | 13.18 | 0.00 | |
| 1003A | 84-QAM Waveform, 100 MHz | X | 2.63 | 68.51 | 18.25 | 3.91 | 150.0 $\pm 1.0\%$ $\pm 0.8\%$ |
| 1003A | AAA | Y | 1.74 | 64.25 | 13.18 | 0.00 | |
| 1003A | AAA | Z | 1.75 | 64.72 | 13.99 | 0.00 | |
| 1003A | 84-QAM Waveform, 40 MHz | X | 3.34 | 66.30 | 15.25 | 0.00 | 150.0 $\pm 2.0\%$ $\pm 0.8\%$ |
| 1003A | AAA | Y | 1.74 | 64.25 | 13.18 | 0.00 | |
| 1003A | AAA | Z | 2.70 | 65.72 | 14.74 | 0.00 | |
| 1003A | WLAN DCCOF, 84-QAM, 40MHz | X | 1.11 | 65.35 | 13.77 | 0.00 | 150.0 $\pm 3.0\%$ $\pm 0.8\%$ |
| 1003A | AAA | Y | 4.70 | 65.54 | 15.41 | 0.00 | |
| 1003A | AAA | Z | 3.83 | 66.16 | 15.28 | 0.00 | |

Note: For details on UID parameters see Appendix

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

* The uncertainties of Norm. K, Z do not affect the E1 field uncertainty value. (See Page 3 and 5)
 * Numerical transition parameter - uncertainty not required.
 * Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No: EX3-7346_Mar22 Page 3 of 24

EX3DV4 - SN:7346 March 30, 2022

DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346

Sensor Model Parameters

| ST | C2 | IP | IP | T1 | T2 | T3 | T4 | T5 | T6 |
|----|------|--------|-------|------|------|------|------|------|------|
| X | 16.5 | 291.80 | 26.10 | 5.61 | 0.00 | 0.00 | 1.42 | 5.12 | 1.01 |
| Y | 37.1 | 270.84 | 34.12 | 8.29 | 0.00 | 0.00 | 5.01 | 1.82 | 0.09 |
| Z | 9.7 | 69.74 | 33.37 | 4.96 | 0.00 | 0.00 | 4.94 | 0.61 | 0.00 |

Other Probe Parameters

| | |
|---|------------|
| Sensor Arrangement | Triangular |
| Connector Angle (°) | -168.1 |
| Mechanical Surface Detection Mode | enabled |
| Optical Surface Detection Mode | disabled |
| Probe Contact Length | 337 mm |
| Probe Body Diameter | 10 mm |
| Tip Length | 9 mm |
| Tip Diameter | 2.5 mm |
| Probe Tip to Sensor X Calibration Point | 1 mm |
| Probe Tip to Sensor Y Calibration Point | 1 mm |
| Probe Tip to Sensor Z Calibration Point | 1 mm |
| Recommended Measurement Distance from Surface | 1.4 mm |

Note: Measurement distance from surface can be increased to 3-4 mm for an Area Scan job.

Certificate No: EX3-7346_Mar22 Page 4 of 24

EX3DV4 - SN:7346 March 30, 2022

DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346

Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) | Relative Permittivity ¹ | Conductivity (S/m) ² | Const. X | Const. Y | Const. Z | Alpha ³ | Depth ⁴ (mm) | Unc. (k=2) |
|---------|------------------------------------|---------------------------------|----------|----------|----------|--------------------|-------------------------|--------------|
| 750 | 41.9 | 0.89 | 10.56 | 10.56 | 10.56 | 0.55 | 0.85 | $\pm 12.0\%$ |
| 815 | 41.6 | 0.90 | 10.12 | 10.12 | 10.12 | 0.42 | 0.96 | $\pm 12.0\%$ |
| 900 | 41.6 | 0.97 | 10.10 | 10.10 | 10.10 | 0.53 | 0.80 | $\pm 12.0\%$ |
| 1450 | 40.5 | 1.20 | 9.26 | 9.26 | 9.26 | 0.50 | 0.80 | $\pm 12.0\%$ |
| 1750 | 40.1 | 1.37 | 8.83 | 8.83 | 8.83 | 0.34 | 0.86 | $\pm 12.0\%$ |
| 1900 | 40.2 | 1.40 | 8.48 | 8.48 | 8.48 | 0.35 | 0.86 | $\pm 12.0\%$ |
| 2000 | 40.0 | 1.40 | 8.35 | 8.35 | 8.35 | 0.34 | 0.86 | $\pm 12.0\%$ |
| 2300 | 39.5 | 1.67 | 7.86 | 7.86 | 7.86 | 0.39 | 0.90 | $\pm 12.0\%$ |
| 2450 | 39.2 | 1.80 | 7.63 | 7.63 | 7.63 | 0.41 | 0.90 | $\pm 12.0\%$ |
| 2800 | 39.0 | 1.96 | 7.33 | 7.33 | 7.33 | 0.44 | 0.90 | $\pm 12.0\%$ |
| 3300 | 38.2 | 2.71 | 7.15 | 7.15 | 7.15 | 0.30 | 1.35 | $\pm 13.1\%$ |
| 3500 | 37.9 | 2.91 | 7.14 | 7.14 | 7.14 | 0.30 | 1.35 | $\pm 13.1\%$ |
| 3750 | 37.7 | 3.12 | 6.85 | 6.85 | 6.85 | 0.30 | 1.35 | $\pm 13.1\%$ |
| 3900 | 37.5 | 3.32 | 6.71 | 6.71 | 6.71 | 0.40 | 1.60 | $\pm 13.1\%$ |
| 4100 | 37.2 | 3.53 | 6.58 | 6.58 | 6.58 | 0.40 | 1.60 | $\pm 13.1\%$ |
| 4200 | 37.1 | 3.63 | 6.30 | 6.30 | 6.30 | 0.40 | 1.70 | $\pm 13.1\%$ |
| 4400 | 36.9 | 3.84 | 6.24 | 6.24 | 6.24 | 0.40 | 1.70 | $\pm 13.1\%$ |
| 4600 | 36.7 | 4.04 | 6.11 | 6.11 | 6.11 | 0.40 | 1.70 | $\pm 13.1\%$ |
| 4800 | 36.4 | 4.25 | 6.08 | 6.08 | 6.08 | 0.40 | 1.80 | $\pm 13.1\%$ |
| 4950 | 36.3 | 4.40 | 5.84 | 5.84 | 5.84 | 0.40 | 1.80 | $\pm 13.1\%$ |
| 5200 | 36.0 | 4.66 | 5.25 | 5.25 | 5.25 | 0.40 | 1.80 | $\pm 13.1\%$ |
| 5300 | 35.9 | 4.70 | 5.12 | 5.12 | 5.12 | 0.40 | 1.80 | $\pm 13.1\%$ |
| 5500 | 35.6 | 4.96 | 4.85 | 4.85 | 4.85 | 0.40 | 1.80 | $\pm 13.1\%$ |
| 5600 | 35.5 | 5.07 | 4.70 | 4.70 | 4.70 | 0.40 | 1.80 | $\pm 13.1\%$ |
| 5800 | 35.3 | 5.27 | 4.75 | 4.75 | 4.75 | 0.40 | 1.80 | $\pm 13.1\%$ |

* Frequency validity above 300 MHz or a 100 MHz only applies for DASY v4 and higher (see Page 2). Also it is restricted to 100 MHz. The uncertainty is the RSS of the Const. uncertainty at calibration frequency and the uncertainty for the measured frequency used. Frequency validity below 300 MHz is 10, 20, 40, 60 and 70 MHz for Const. assessments at 20, 40, 100, 150 and 200 MHz respectively. Validity of Const. assessed at 8 MHz is 6 MHz, and Const. assessed at 10 MHz is 10 MHz. Only Frequency validity can be extended to 10 MHz.
 * All frequencies below 3 GHz, the validity of tissue parameters is and can be treated as a 10% liquid compensation formula is applied to measured field values. All frequencies above 3 GHz, the validity of tissue parameters is and can be restricted to 5%. The uncertainty is the RSS of the Const. uncertainty for indicated target tissue parameters.
 * Alpha/Depth are determined during calibration. SPICAC warrants that the remaining deviation due to the boundary effect after compensation is always less than 1% for frequencies below 3 GHz and below 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

Certificate No: EX3-7346_Mar22 Page 5 of 24

EX3DV4 - SN:7346 March 30, 2022

DASY/EASY - Parameters of Probe: EX3DV4 - SN:7346

Calibration Parameter Determined in Head Tissue Simulating Media

| f (MHz) | Relative Permittivity ¹ | Conductivity (S/m) ² | Const. X | Const. Y | Const. Z | Alpha ³ | Depth ⁴ (mm) | Unc. (k=2) |
|---------|------------------------------------|---------------------------------|----------|----------|----------|--------------------|-------------------------|--------------|
| 6900 | 34.5 | 6.07 | 5.30 | 5.30 | 5.30 | 0.20 | 2.50 | $\pm 18.6\%$ |


* Frequency validity above 3 GHz is 700 MHz. The uncertainty is the RSS of the Const. uncertainty at calibration frequency and the uncertainty for the measured frequency used.
 * All frequencies 6-10 GHz, the validity of tissue parameters is and can be treated as a 10% liquid compensation formula is applied to measured field values. The uncertainty is the RSS of the Const. uncertainty for indicated target tissue parameters.
 * Alpha/Depth are determined during calibration. SPICAC warrants that the remaining deviation due to the boundary effect after compensation is always less than 1% for frequencies below 3 GHz, below 2% for frequencies between 3-6 GHz, and below 4% for frequencies between 6-10 GHz at any distance larger than half the probe tip diameter from the boundary.

Certificate No: EX3-7346_Mar22 Page 6 of 24



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com
 No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300

(86-512)57355888 (86-512)57370818 www.sgs.com.cn
 (86-512)57355888 (86-512)57370818 sgs.china@sgs.com



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN_Doccheck@sgs.com
 No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
 中国·江苏·昆山市留学院创业园伟业路10号 邮编 215300



Table with columns: Appendix: Communication Calibration Parameters, LNO, Rev, Manufacturer, System Name, Group, Pass, Unit, and Pass Rate. It contains two main sections of data, one on the left and one on the right, each with a table of test results and a corresponding certificate reference.



Unless otherwise agreed in writing, this document is issued by its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/terms-and-conditions.aspx and, for electronic form documents subject to Terms and Conditions for Electronic Documents at http://www.sgs.com/en/terms-and-conditions/terms-e-document.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced or used in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only. Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300 t(86-512)57355888 t(86-512)57370818 www.sgs.com.cn 中国·江苏·昆山学生创业园伟业路10号 邮编 215300 t(86-512)57355888 t(86-512)57370818 sgs.china@sgs.com



Table with columns: Reference No., Part No., Description, Test Method, Result, and Date. Includes test results for various electronic components.

Table with columns: Reference No., Part No., Description, Test Method, Result, and Date. Includes test results for various electronic components.

Table with columns: Reference No., Part No., Description, Test Method, Result, and Date. Includes test results for various electronic components.

Table with columns: Reference No., Part No., Description, Test Method, Result, and Date. Includes test results for various electronic components.



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed on request or available at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced in full, without prior written approval of the Company. Any unauthorized translation, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only. Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300 (86-512)57355888 (86-512)57370818 www.sgsngs.com.cn 中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300 (86-512)57355888 (86-512)57370818 sgs.china@sgs.com

| EX3D04 - SN 7346 | | | | March 30, 2022 | | | |
|------------------|-----|-------------------------------------|------|----------------|---------|--|--|
| 10673 | AAC | IEEE 802.11a (20MHz) MCS2 (900-00) | WLAN | 8.78 | ± 0.6 % | | |
| 10674 | AAC | IEEE 802.11a (20MHz) MCS3 (900-00) | WLAN | 8.76 | ± 0.6 % | | |
| 10675 | AAC | IEEE 802.11a (20MHz) MCS4 (900-00) | WLAN | 8.80 | ± 0.6 % | | |
| 10676 | AAC | IEEE 802.11a (20MHz) MCS5 (900-00) | WLAN | 8.77 | ± 0.6 % | | |
| 10677 | AAC | IEEE 802.11a (20MHz) MCS6 (900-00) | WLAN | 8.73 | ± 0.6 % | | |
| 10678 | AAC | IEEE 802.11a (20MHz) MCS7 (900-00) | WLAN | 8.78 | ± 0.6 % | | |
| 10679 | AAC | IEEE 802.11a (20MHz) MCS8 (900-00) | WLAN | 8.80 | ± 0.6 % | | |
| 10680 | AAC | IEEE 802.11a (20MHz) MCS9 (900-00) | WLAN | 8.80 | ± 0.6 % | | |
| 10681 | AAC | IEEE 802.11a (20MHz) MCS10 (900-00) | WLAN | 8.82 | ± 0.6 % | | |
| 10682 | AAC | IEEE 802.11a (20MHz) MCS11 (900-00) | WLAN | 8.83 | ± 0.6 % | | |
| 10683 | AAC | IEEE 802.11a (20MHz) MCS12 (900-00) | WLAN | 8.82 | ± 0.6 % | | |
| 10684 | AAC | IEEE 802.11a (20MHz) MCS13 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10685 | AAC | IEEE 802.11a (20MHz) MCS14 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10686 | AAC | IEEE 802.11a (20MHz) MCS15 (900-00) | WLAN | 8.78 | ± 0.6 % | | |
| 10687 | AAC | IEEE 802.11a (20MHz) MCS16 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10688 | AAC | IEEE 802.11a (20MHz) MCS17 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10689 | AAC | IEEE 802.11a (20MHz) MCS18 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10690 | AAC | IEEE 802.11a (20MHz) MCS19 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10691 | AAC | IEEE 802.11a (20MHz) MCS20 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10692 | AAC | IEEE 802.11a (20MHz) MCS21 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10693 | AAC | IEEE 802.11a (20MHz) MCS22 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10694 | AAC | IEEE 802.11a (20MHz) MCS23 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10695 | AAC | IEEE 802.11a (20MHz) MCS24 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10696 | AAC | IEEE 802.11a (20MHz) MCS25 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10697 | AAC | IEEE 802.11a (20MHz) MCS26 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10698 | AAC | IEEE 802.11a (20MHz) MCS27 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10699 | AAC | IEEE 802.11a (20MHz) MCS28 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10700 | AAC | IEEE 802.11a (20MHz) MCS29 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10701 | AAC | IEEE 802.11a (20MHz) MCS30 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10702 | AAC | IEEE 802.11a (20MHz) MCS31 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10703 | AAC | IEEE 802.11a (20MHz) MCS32 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10704 | AAC | IEEE 802.11a (20MHz) MCS33 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10705 | AAC | IEEE 802.11a (20MHz) MCS34 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10706 | AAC | IEEE 802.11a (20MHz) MCS35 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10707 | AAC | IEEE 802.11a (20MHz) MCS36 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10708 | AAC | IEEE 802.11a (20MHz) MCS37 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10709 | AAC | IEEE 802.11a (20MHz) MCS38 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10710 | AAC | IEEE 802.11a (20MHz) MCS39 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10711 | AAC | IEEE 802.11a (20MHz) MCS40 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10712 | AAC | IEEE 802.11a (20MHz) MCS41 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10713 | AAC | IEEE 802.11a (20MHz) MCS42 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10714 | AAC | IEEE 802.11a (20MHz) MCS43 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10715 | AAC | IEEE 802.11a (20MHz) MCS44 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10716 | AAC | IEEE 802.11a (20MHz) MCS45 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10717 | AAC | IEEE 802.11a (20MHz) MCS46 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10718 | AAC | IEEE 802.11a (20MHz) MCS47 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10719 | AAC | IEEE 802.11a (20MHz) MCS48 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10720 | AAC | IEEE 802.11a (20MHz) MCS49 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10721 | AAC | IEEE 802.11a (20MHz) MCS50 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10722 | AAC | IEEE 802.11a (20MHz) MCS51 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10723 | AAC | IEEE 802.11a (20MHz) MCS52 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10724 | AAC | IEEE 802.11a (20MHz) MCS53 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10725 | AAC | IEEE 802.11a (20MHz) MCS54 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10726 | AAC | IEEE 802.11a (20MHz) MCS55 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10727 | AAC | IEEE 802.11a (20MHz) MCS56 (900-00) | WLAN | 8.79 | ± 0.6 % | | |
| 10728 | AAC | IEEE 802.11a (20MHz) MCS57 (900-00) | WLAN | 8.79 | ± 0.6 % | | |

Certificate No. EX3-7346_M022 Page 19 of 24

Certificate No. EX3-7346_M022 Page 20 of 24



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.
Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com
No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300
中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300

| | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
|-------|--|--|---------------|--|---------------|------|--------|-------|-----|--|---------------|------|--------|-------|-----|--|---------------|------|--------|-------|-----|--|---------------|------|--------|-------|-----|--|---------------|------|--------|-------|-----|--|---------------|------|--------|
| | <p>EX32V4-SN 7346 March 30, 2022</p> <table border="1"> <tr> <td>13985</td> <td>AAA</td> <td>SG NR DL (CP-QPDM, TM 3.1, 40 MHz, 64-GAM, 30 kHz)</td> <td>SG NR FR1 T0D</td> <td>9.54</td> <td>± 9.5%</td> </tr> <tr> <td>13986</td> <td>AAA</td> <td>SG NR DL (CP-QPDM, TM 3.1, 50 MHz, 64-GAM, 30 kHz)</td> <td>SG NR FR1 T0D</td> <td>9.92</td> <td>± 9.5%</td> </tr> <tr> <td>13987</td> <td>AAA</td> <td>SG NR DL (CP-QPDM, TM 3.1, 60 MHz, 64-GAM, 30 kHz)</td> <td>SG NR FR1 T0D</td> <td>9.93</td> <td>± 9.5%</td> </tr> <tr> <td>13988</td> <td>AAA</td> <td>SG NR DL (CP-QPDM, TM 3.1, 70 MHz, 64-GAM, 30 kHz)</td> <td>SG NR FR1 T0D</td> <td>9.38</td> <td>± 9.5%</td> </tr> <tr> <td>13989</td> <td>AAA</td> <td>SG NR DL (CP-QPDM, TM 3.1, 80 MHz, 64-GAM, 30 kHz)</td> <td>SG NR FR1 T0D</td> <td>9.23</td> <td>± 9.5%</td> </tr> <tr> <td>13990</td> <td>AAA</td> <td>SG NR DL (CP-QPDM, TM 3.1, 90 MHz, 64-GAM, 30 kHz)</td> <td>SG NR FR1 T0D</td> <td>9.52</td> <td>± 9.5%</td> </tr> </table> <p><small>* Linearity is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.</small></p> <p>Certificate No: EX3-7346_Mar22 Page 24 of 24</p> | 13985 | AAA | SG NR DL (CP-QPDM, TM 3.1, 40 MHz, 64-GAM, 30 kHz) | SG NR FR1 T0D | 9.54 | ± 9.5% | 13986 | AAA | SG NR DL (CP-QPDM, TM 3.1, 50 MHz, 64-GAM, 30 kHz) | SG NR FR1 T0D | 9.92 | ± 9.5% | 13987 | AAA | SG NR DL (CP-QPDM, TM 3.1, 60 MHz, 64-GAM, 30 kHz) | SG NR FR1 T0D | 9.93 | ± 9.5% | 13988 | AAA | SG NR DL (CP-QPDM, TM 3.1, 70 MHz, 64-GAM, 30 kHz) | SG NR FR1 T0D | 9.38 | ± 9.5% | 13989 | AAA | SG NR DL (CP-QPDM, TM 3.1, 80 MHz, 64-GAM, 30 kHz) | SG NR FR1 T0D | 9.23 | ± 9.5% | 13990 | AAA | SG NR DL (CP-QPDM, TM 3.1, 90 MHz, 64-GAM, 30 kHz) | SG NR FR1 T0D | 9.52 | ± 9.5% |
| 13985 | AAA | SG NR DL (CP-QPDM, TM 3.1, 40 MHz, 64-GAM, 30 kHz) | SG NR FR1 T0D | 9.54 | ± 9.5% | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 13986 | AAA | SG NR DL (CP-QPDM, TM 3.1, 50 MHz, 64-GAM, 30 kHz) | SG NR FR1 T0D | 9.92 | ± 9.5% | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 13987 | AAA | SG NR DL (CP-QPDM, TM 3.1, 60 MHz, 64-GAM, 30 kHz) | SG NR FR1 T0D | 9.93 | ± 9.5% | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 13988 | AAA | SG NR DL (CP-QPDM, TM 3.1, 70 MHz, 64-GAM, 30 kHz) | SG NR FR1 T0D | 9.38 | ± 9.5% | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 13989 | AAA | SG NR DL (CP-QPDM, TM 3.1, 80 MHz, 64-GAM, 30 kHz) | SG NR FR1 T0D | 9.23 | ± 9.5% | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| 13990 | AAA | SG NR DL (CP-QPDM, TM 3.1, 90 MHz, 64-GAM, 30 kHz) | SG NR FR1 T0D | 9.52 | ± 9.5% | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |

4 Impedance and return loss

| Dipole CLA150 SN 4025 | | | | |
|-----------------------|-----------------|-----|---------------|----|
| Head Liquid | | | | |
| Date of Measurement | Return Loss(dB) | Δ % | Impedance (Ω) | ΔΩ |
| 2021/4/26 | -31.4 | / | 47.8 | / |
| Dipole D450V3 SN 1103 | | | | |
| Head Liquid | | | | |
| Date of Measurement | Return Loss(dB) | Δ % | Impedance (Ω) | ΔΩ |
| 2021/4/21 | -23 | / | 57.1 | / |



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.sgs.com/en/Terms-and-Conditions_2828x and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: CN.Doccheck@sgs.com

No.10, Weiye Road, Innovation Park, Kunshan, Jiangsu, China 215300 t(86-512)57355888 f(86-512)57370818 www.sgsgroup.com.cn
 中国·江苏·昆山市留学生创业园伟业路10号 邮编 215300 t(86-512)57355888 f(86-512)57370818 sgs.china@sgs.com