



FCC Class II Permissive Change Request Letter

Date: August 29, 2024

To Federal Communications Commission:

RE: FCC Class II Permissive Change Request for Company: Stryker Medical **FCC ID:** Z7A-SDMACP

We are submitting an application for a class II permissive change to the FCC for approval of the Company name: Stryker Medical, product description: Embedded Wireless Module (FCC ID: Z7A-SDMACP, Original Grant Date: 09/15/2020). The transmitter module itself has not changed. The antenna and rated power have also not changed. Here are the changes:

Collocation with Bluetooth transmitter FCC ID: Z7ALBCA1KU1WA. Note the Bluetooth transmitter for FCC ID: Z7A-SDMACP is disabled while used with Bluetooth transmitter FCC ID: Z7ALBCA1KU1WA.

Sincerely,

A handwritten signature in black ink that reads "Divya Murali".

Divya Murali

Director, Global Regulatory Affairs

Divya.Murali@stryker.com

Medical

3800 E. Centre Avenue, Portage, MI 49002 USA
P +1 269 389 6998