NCL CALIBRATION LABORATORIES

Calibration File No.: PC-1419

Client.: ISL

CERTIFICATE OF CALIBRATION

It is certified that the equipment identified below has been calibrated in the **NCL CALIBRATION LABORATORIES** by qualified personnel following recognized procedures and using transfer standards traceable to NRC/NIST.

Equipment: Miniature Isotropic RF Probe Record of Calibration Body

Manufacturer: APREL Laboratories

Model No.: E-020

Serial No.: 266

Calibration Procedure: D01-032-E020-V2, D22-012-Tissue, D28-002-Dipole

Project No: ISL-E020-5683

Calibrated: 20th August 2012 Released on: 20th August 2012

This Calibration Certificate is Incomplete Unless Accompanied with the Calibration Results Summary

Released By:

Art Brennan, Quality Manager

VCL CALIBRATION LABORAT

303 Terry Fox Drive, Suite 102 Kanata, Ontario CANADA K2K 3J1 Division of APREL TEL: (613) 435-8300 FAX: (613) 435-8306

Introduction

This Calibration Report reproduces the results of the calibration performed in line with the references listed below. Calibration is performed using accepted methodologies as per the references listed below. Probes are calibrated for air, and tissue and the values reported are the results from the physical quantification of the probe through meteorgical practices.

Calibration Method

Probes are calibrated using the following methods.

<1000MHz

TEM Cell for sensitivity in air

Standard phantom using temperature transfer method for sensitivity in tissue

>1000MHz

Waveguide* method to determine sensitivity in air and tissue

*Waveguide is numerically (simulation) assessed to determine the field distribution and power

The boundary effect for the probe is assessed using a standard flat phantom where the probe output is compared against a numerically simulated series of data points

References

- IEEE Standard 1528 (2003) including Amendment 1
 IEEE Recommended Practice for Determining the Peak Spatial-Average Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques
- o EN 62209-1 (2006)
 - Human Exposure to RF Fields from hand-held and body-mounted wireless communication devices Human models. instrumentation, and procedures-Part 1: Procedure to measure the Specific Absorption Rate (SAR) for hand-held mobile wireless devices
- IEC 62209-2 Ed. 1.0 (2010-03)
 Human exposure to RF fields from hand-held and body-mounted wireless devices Human models, instrumentation, and procedures Part 2: specific absorption rate (SAR) for wireless communication devices (30 MHz 6 GHz)
- o TP-D01-032-E020-V2 E-Field probe calibration procedure
- o D22-012-Tissue dielectric tissue calibration procedure
- D28-002-Dipole procedure for validation of SAR system using a dipole
- IEEE 1309 Draft Standard for Calibration of Electromagnetic Field Sensors and Probes,
 Excluding Antennas, from 9kHz to 40GHz

Conditions

Probe 266 was a recalibration.

Ambient Temperature of the Laboratory: $22 \,^{\circ}\text{C} \, +/- \, 1.5 \,^{\circ}\text{C}$ Temperature of the Tissue: $21 \,^{\circ}\text{C} \, +/- \, 1.5 \,^{\circ}\text{C}$

Relative Humidity: < 60%

Primary Measurement Standards

Instrument	Serial Number	Cal due date
Power meter Anritsu MA2408A	90025437	Nov.4, 2012
Power Sensor Anritsu MA2481D	103555	Nov 4, 2012
Attenuator HP 8495A (70dB)	1944A10711	Sept. 14, 2012
Network Analyzer Anritsu MT8801C	MB11855	Feb. 8, 2013

Secondary Measurement Standards

Signal Generator Agilent E4438C -506 MY55182336 June 7, 2013

Attestation

The below named signatories have conducted the calibration and review of the data which is presented in this calibration report.

We the undersigned attest that to the best of our knowledge the calibration of this subject has been accurately conducted and that all information contained within the results pages have been reviewed for accuracy.

Art Brennan, Quality Manager

Dan Brooks, Test Engineer

NCL Calibration Laboratories

Division of APREL Inc.

Probe Summary

Probe Type: E-Field Probe E020

Serial Number: 266

Frequency: As presented on page 5

Sensor Offset: 1.56

Sensor Length: 2.5

Tip Enclosure: Composite*

Tip Diameter: < 2.9 mm

Tip Length: 55 mm

Total Length: 289 mm

Sensitivity in Air

Diode Compression Point: 95 mV

^{*}Resistive to recommended tissue recipes per IEEE-1528

Calibration for Tissue (Head H, Body B)

Frequency	Tissue Type	Measured Epsilon	Measured Sigma	Calibration Uncertainty	Tolerance Uncertainty for 5%*	Conversion Factor
450 H	Head	Х	Х	Х	Х	Х
450 B	Body	Х	Х	X	Х	Х
750 H	Head	Х	Х	Х	Х	Х
750 B	Body	Х	Х	Х	Х	Х
835 H	Head	Х	Х	Х	Х	Х
835 B	Body	<mark>56.31</mark>	<mark>0.95</mark>	<mark>3.5</mark>	<mark>3.4</mark>	<mark>6.8</mark>
900 H	Head	X	X	Х	Х	X
900 B	Body	<mark>56.08</mark>	1.05	<mark>3.5</mark>	3.4	<mark>6.7</mark>
1450 H	Head	X	Х	Х	Х	Х
1450 B	Body	Х	Х	Х	Х	Х
1500 H	Head	Х	Х	Х	Х	Х
1500 B	Body	X	Х	Х	Х	Х
1640 H	Head	X	Х	Х	Х	X
1640 B	Body	X	Х	Х	Х	Х
1750 H	Head	X	Х	Х	Х	X
1750 B	Body	X	Х	X	Х	Х
1800 H	Head	X	Х	Х	Х	X
1800 B	Body	<mark>52.16</mark>	<mark>1.56</mark>	<mark>3.5</mark>	<mark>2.7</mark>	<mark>5.5</mark>
1900 H	Head	X	Х	Х	Х	X
1900 B	<mark>Body</mark>	<mark>51.57</mark>	<mark>1.57</mark>	<mark>3.5</mark>	2.7	<mark>5.4</mark>
2000 H	Head	Х	Х	X	X	Х
2000 B	Body	Х	Х	X	X	Х
2100 H	Head	Х	Х	X	X	X
2100 B	Body	Х	Х	X	X	X
2300 H	Head	Х	Х	X	X	Х
2300 B	Body	Х	Х	Х	Х	Х
2450 H	Head	Х	Х	X	X	X
2450B	<mark>Body</mark>	50.22	<mark>1.93</mark>	3.5	3.5	<mark>4.55</mark>
2600 H	Head	X	X	X	X	X
2600 B	Body	X	X	X	X	X
3000 H	Head	X	X	X	X	X
3000 B	Body	Х	X	X	X	X
3600 H	Head	Х	Х	Х	Х	Х
3600 B	Body	Х	Х	Х	Х	Х
5200 H	Head	Х	Х	Х	Х	Х
5200 B	<mark>Body</mark>	46.84	<mark>5.08</mark>	<mark>3.5</mark>	<mark>3.5</mark>	<mark>3.3</mark>
5600 H	Head	Х	Х	Х	Х	Х
5600 B	<mark>Body</mark>	45.83	<mark>5.63</mark>	<mark>3.5</mark>	<mark>3.5</mark>	3
5800 H	Head	Х	Х	X	X	Х
5800 B	<mark>Body</mark>	<mark>45.87</mark>	<mark>5.93</mark>	<mark>3.5</mark>	3.5	3.4

Page 5 of 10

This page has been reviewed for content and attested to on Page 2 of this document.

NCL Calibration Laboratories

Division of APREL Inc.

Boundary Effect:

Uncertainty resulting from the boundary effect is less than 2.1% for the distance between the tip of the probe and the tissue boundary, when less than 0.58mm.

Spatial Resolution:

The spatial resolution uncertainty is less than 1.5% for 4.9mm diameter probe.

The spatial resolution uncertainty is less than 1.0% for 2.5mm diameter probe.

DAQ-PAQ Contribution

To minimize the uncertainty calculation all tissue sensitivity values were calculated using a load impedance of 5 M Ω .


Boundary Effect:

For a distance of 0.58mm the worst case evaluated uncertainty (increase in the probe sensitivity) is less than 2.1%.


NOTES:


*The maximum deviation from the centre frequency when comparing the lower to upper range is listed.

Receiving Pattern Air



Isotropy Error Air







Isotropicity Tissue:


0.10 dB

Dynamic Range



Video Bandwidth





Video Bandwidth at 500 Hz 1 dB Video Bandwidth at 1.02 KHz: 3 dB

Test Equipment

The test equipment used during Probe Calibration, manufacturer, model number and, current calibration status are listed and located on the main APREL server R:\NCL\Calibration Equipment\Instrument List May 2011.