

Shenzhen Mindray Bio-Medical Electronics Co., Ltd.

Federal Communications Commission
Equipment Authorization Division
7435 Oakland Mills Road
Columbia, MD 21046
USA

Date: 2023-3-14

Subject; wireless module

Company name: Shenzhen Mindray Bio-Medical Electronics Co., Ltd.
FCC ID: ZLZ-WL6PR

Dear leo Shao,

This letter includes the FCC application requirements for Modular Transmitter Approval Request for;-

FCC KDB 996369 D01 'Module Certification Guide v02'; and
FCC KDB 996369 D03 OEM Manual v01

In accordance with 47CFR 15.212 Modular Transmitters and KDB 996369 D01 'Module Equip Auth Guide v02'. FCC ID ZLZ-WL6PR has been examined against the following requirements.

Requirement per 15.212 and KDB 996369 D01	Explanation from Grantee (do not write yes/no, but explain why product complies/how it is achieved)
The radio elements must have the radio frequency circuitry shielded. Physical components and tuning capacitor(s) may be located external to the shield, but must be on the module assembly.	Yes, the module has the radio frequency circuitry shielded.
The module must have buffered modulation/data inputs to ensure that the device will comply with Part 15 requirements with any type of input signal.	Yes, all inputs to the modules are buffered through logic or microprocessor inputs.
The module must contain power supply regulation on the module.	Yes, the module contains a power supply regulation on the module.
The module must contain a permanently attached antenna, or contain a unique antenna connector, and be marketed and operated only with specific antenna(s), per §§ 15.203, 15.204(b), 15.204(c), 15.212(a), 2.929(b).	The module ships with six types of antennas.
The module must demonstrate compliance in a stand-alone configuration.	Yes, the module demonstrates compliance in a stand-alone configuration.
The module must be labeled with its permanently affixed FCC ID label, or use an electronic display (see KDB Publication 784748).	Yes, the module has its permanently affixed FCC ID label and the FCC ID label is printed on the top of the shield.
The module must comply with all specific rules applicable to the transmitter, including all the conditions provided in the integration instructions by the grantee.	Yes, the module complies with all specific rules applicable to the transmitter including all the conditions provided in the integration instructions by the grantee, refer to test report and user manual.

Shenzhen Mindray Bio-Medical Electronics Co., Ltd.

The module must comply with RF exposure requirements	Yes, transmitter meets MPE calculation of 47 CFR 1.1307 and Part 2.1091. Refer to MPE Reports and Refer to modular installation manual
--	--

Integration Instructions for host product manufacturers

The following items are submitted in support of application for Modular Transmitter FCC ID as noted above as required by the FCC KDB 996369 D03 OEM Manual v01.

These items are provided as integration instructions for host product manufacturers (e.g., OEM instruction manual) to use when integrating a module in a host product.

Any requirements that are not applicable to the Module are as indicated below.

Summary of requirements and Checklist. Refer to the KDB for description of the complete requirements;

KDB Ref Sect	Requirements of KDB 996369 D03	User Manual Page Number reference
2.2	List of applicable FCC rules	Page 10
2.3	Summarize the specific operational use conditions	Page 11
2.4	Limited module procedures	Page 11
2.5	Trace antenna designs	Page 12
2.6	RF exposure considerations	Page 13
2.7	Antennas	Page 13
2.8	Label and compliance information	Page 14
2.9	Information on test modes and additional testing requirements	Page 14
2.10	Additional testing, Part 15 Subpart B disclaimer	Page 14

Name: leo Shao

Date: 2023-3-14

Title: Manager

Signature of applicant

Leo Shao