

Shenzhen Mindray BIO-Medical electronics Co., LTD.
Mindray Building, Keji 12th Road South, Hi-tech Ind, Shenzhen, China

FCC Software Configuration Control Declaration

Date: September 14, 2024

Federal Communications Commission
7435 Oakland Mills Road
Columbia, MD 21046
USA

Attn: OET Dept.

Product/ Model Number: WiFi module/WXT5CM2803

FCC ID: ZLZ-WXT5CM2803

Shenzhen Mindray BIO-Medical electronics Co., LTD. (“the applicant”) hereby attest that this device complies with the following requirements of Part 15E of the FCC's rules for the 6GHz bands:

a.) 1. Contention-Based Protocol, as demonstrated in the test report, is permanently embedded in the module and is not host-dependent.


2. That the device will only associate and connect with a low-power indoor access point or subordinate device and never directly connect to other client devices.

3. That this device will always initiate transmission under the control of a low-power indoor AP or subordinate except for brief transmissions before joining a network. These short messages will only occur if the client has detected an indoor AP or subordinate operating on a channel. These brief messages will have a time-out mechanism such that if it does not receive a response from an AP it will not continually repeat the request.

4. That transmission will be lower or equal to the power advertised by the indoor low-power access point or subordinate and never above the maximum output power allowed by the FCC grant for equipment class 6XD.

b.) The statement acknowledging device restrictions:

Prohibited for control of or communications with unmanned aircraft systems, including drones.

Signature: 

Name: Luo jihao

Title: Manager

Company: Shenzhen Mindray BIO-Medical electronics Co., LTD.

Address: Mindray Building, Keji 12th Road South, Hi-tech Ind, Shenzhen, China

Date: 2023.04.03

3.6 GHz attestation required 0914

3.6 GHz attestation required 0914